



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94587

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.4	4.01	04/12/2018	12:10
2	CAL2	1	Water			20.4	7.00	04/12/2018	12:11
3	CAL3	1	Water			20.3	10.03	04/12/2018	12:13
4	ICV	1	Water			20.3	7.01	04/12/2018	12:15
5	CCV1	1	Water			20.2	2.01	04/12/2018	12:16
6	J2338-01	1	Solid	20.02	20	21.4	7.86	04/12/2018	12:19
7	J2338-01DUP	1	Solid	20.04	20	21.5	7.87	04/12/2018	12:20
8	CCV2	1	Water			20.3	12.01	04/12/2018	12:21

LB 94587

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph J2338

WorkList ID : 110703

Date : 4/12/2018 11:55:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/29/2018	Solid	J2338-01	pH	Cool 4 deg C	ENTA04	N12	E2	03/29/2018	9045C

Date/Time 04.12.18 12.00 PM
Received by: Jn
Relinquished by: Cn

Date/Time 04.12.18 12.35 PM
Received by: CD
Relinquished by: JF