



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94593

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	04/13/2018	08:45
2	J2336-02		1	>100.0	>212.0	>212.0	04/13/2018	09:20
3	J2336-02DUP		1	>100.0	>212.0	>212.0	04/13/2018	09:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB94593

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-041318

WorkList ID : 110748

Date : 4/13/2018 8:11:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2018	Water	J2336-02	C Flash Point	Cool 4 deg C	PSEG01	N63	SL	04/11/2018	1010A

Date/Time 4/13/18 8:15 AM
Received by: KS
Relinquished by: CP

Date/Time 4/13/18 09:55 AM
Received by: CP
Relinquished by: KS