

2194594

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/13/2018 10:53

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.022	0.0	0.107	
ICB1	0.009	0.0	0.004	
CCV1	0.975	0.0	0.102	
CCB1	0.009	0.0	0.004	
PB108200BL	0.005	0.0	0.003	
PB108200BS	1.026	0.0	0.107	
J2318-08	0.014	0.0	0.004	
J2318-08DUP	0.023	0.0	0.005	
J2318-08MS	1.009	0.0	0.106	
J2318-08MSD	1.012	0.0	0.106	
CCV2	0.970	0.0	0.102	
CCB2	-0.003	0.0	0.003	

N	12
Mean	0.506
SD	0.5188
CV%	102.56

Aquakem v. 7.2AQ1

AS

Results from time period:

Fri Apr 13 10:19:05 2018

Fri Apr 13 10:50:50 2018

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.047	mg/l	4/13/2018 10:19:05	
0.1PPM	A	Phenolics-	P	0.1063	mg/l	4/13/2018 10:19:06	
0.25PPM	A	Phenolics-	P	0.2567	mg/l	4/13/2018 10:19:07	
0.50PPM	A	Phenolics-	P	0.4964	mg/l	4/13/2018 10:19:08	
1.0PPM	A	Phenolics-	P	0.9877	mg/l	4/13/2018 10:19:09	
2.0PPM	A	Phenolics-	P	2.006	mg/l	4/13/2018 10:19:10	
ICV1	S	Phenolics-	P	1.0223	mg/l	4/13/2018 10:48:02	
ICB1	S	Phenolics-	P	0.009	mg/l	4/13/2018 10:48:03	
CCV1	S	Phenolics-	P	0.9748	mg/l	4/13/2018 10:48:04	
CCB1	S	Phenolics-	P	0.0087	mg/l	4/13/2018 10:48:05	
PB108200BL	S	Phenolics-	P	0.005	mg/l	4/13/2018 10:48:06	
PB108200BS	S	Phenolics-	P	1.0258	mg/l	4/13/2018 10:48:07	
J2318-08	S	Phenolics-	P	0.0138	mg/l	4/13/2018 10:48:08	
J2318-08DUP	S	Phenolics-	P	0.0226	mg/l	4/13/2018 10:48:09	
J2318-08MS	S	Phenolics-	P	1.0094	mg/l	4/13/2018 10:48:10	
J2318-08MSD	S	Phenolics-	P	1.0121	mg/l	4/13/2018 10:48:11	
CCV2	S	Phenolics-	P	0.97	mg/l	4/13/2018 10:48:12	
CCB2	S	Phenolics-	P	-0.0025	mg/l	4/13/2018 10:48:13	
LCS	S	Phenolics-	P	0.811	mg/l	4/13/2018 10:50:50	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/13/2018 10:25


Test Phenolics-

Accepted 4/13/2018 10:25

Factor 9.842
Bias 0.003

Coeff. of det. 0.999895

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0470	0.0500	
2	Phenol-2	0.014	0.1063	0.1000	
3	Phenol-2	0.029	0.2567	0.2500	
4	Phenol-2	0.053	0.4964	0.5000	
5	Phenol-2	0.103	0.9877	1.0000	
6	Phenol-2	0.207	2.0060	2.0000	