

7470A


L394596

INSTRUMENT ID: CV2

Linear

 μ Abs.:

36075



A= 0.0000e+000

B= 2.7891e-004

C= 1.5167e-002

Rho= 0.9998998

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	31	0.000	31				
0.2	0.200	0.222	0.022	742	0.0 %	742				
2.5	2.500	2.477	-0.023	8825	0.0 %	8825				
5.0	5.000	4.994	-0.006	17852	0.0 %	17852				
7.5	7.500	7.406	-0.094	26499	0.0 %	26499				
10.0	10.000	10.077	0.077	36075	0.0 %	36075				

LB94596 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Apr 2018 11:36:12					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7470A	PPB	13 Apr 2018 11:47:23	S	Std
0.2 - 1	50.2	742	-	0.2000	7470A	PPB	13 Apr 2018 11:49:31	S	Std
2.5 - 1	52.5	8825	-	2.5000	7470A	PPB	13 Apr 2018 11:51:40	S	Std
5.0 - 1	55	17852	-	5.0000	7470A	PPB	13 Apr 2018 11:53:48	S	Std
7.5 - 1	57.5	26499	-	7.5000	7470A	PPB	13 Apr 2018 11:55:56	S	Std
10.0 - 1	510	36075	-	10.0000	7470A	PPB	13 Apr 2018 11:58:04	S	Std
ICV92 - 1	ICV92	13673	3.8288	-	7470A	PPB	13 Apr 2018 12:00:43	U	SMPL
ICB92 - 1	ICB92	-155	-0.0281	-	7470A	PPB	13 Apr 2018 12:02:50	U	SMPL
CCV25 - 1	CCV25	18108	5.0658	-	7470A	PPB	13 Apr 2018 12:04:57	U	SMPL
CCB25 - 1	CCB25	-177	-0.0342	-	7470A	PPB	13 Apr 2018 12:07:03	U	SMPL
CRA - 1	CRA	704	0.2115	-	7470A	PPB	13 Apr 2018 12:09:11	U	SMPL
HighStd - 1	HighStd	36487	10.1919	-	7470A	PPB	13 Apr 2018 12:11:17	U	SMPL
ChkStd - 1	ChkStd	24691	6.9018	-	7470A	PPB	13 Apr 2018 12:13:24	U	SMPL
PB108201BL - 1	PBW	-203	-0.0415	-	7470A	PPB	13 Apr 2018 12:15:34	U	SMPL
PB108201BS - 1	LCSW	12461	3.4907	-	7470A	PPB	13 Apr 2018 12:17:43	U	SMPL
PB108201TB - 1	PB108166TB	-134	-0.0222	-	7470A	PPB	13 Apr 2018 12:19:51	U	SMPL
J2162-02 - 1	EO-01-041118-A	-5	0.0138	-	7470A	PPB	13 Apr 2018 12:21:59	U	SMPL
J2162-02DUP - 1	EO-01-041118-ADUP	0	0.0152	-	7470A	PPB	13 Apr 2018 12:24:08	U	SMPL
J2162-02MS - 1	EO-01-041118-AMS	12704	3.5585	-	7470A	PPB	13 Apr 2018 12:26:17	U	SMPL
J2162-02MSD - 1	EO-01-041118-AMSD	12875	3.6062	-	7470A	PPB	13 Apr 2018 12:28:23	U	SMPL
CCV26 - 1	CCV26	18352	5.1338	-	7470A	PPB	13 Apr 2018 12:30:30	U	SMPL
CCB26 - 1	CCB26	-179	-0.0348	-	7470A	PPB	13 Apr 2018 12:32:38	U	SMPL
J2162-04 - 1	EO-01-041118-B	-8	0.0129	-	7470A	PPB	13 Apr 2018 12:34:46	U	SMPL
J2162-06 - 1	EO-01-041118-C	-18	0.0101	-	7470A	PPB	13 Apr 2018 12:36:53	U	SMPL
J2163-02 - 1	PL-01-041118-A	2	0.0157	-	7470A	PPB	13 Apr 2018 12:38:59	U	SMPL
J2163-04 - 1	PL-01-041118-B	3	0.0160	-	7470A	PPB	13 Apr 2018 12:41:07	U	SMPL
J2163-06 - 1	PL-01-041118-C	1	0.0154	-	7470A	PPB	13 Apr 2018 12:43:14	U	SMPL
J2163-08 - 1	PL-01-041118-D	-19	0.0099	-	7470A	PPB	13 Apr 2018 12:45:22	U	SMPL
J2165-04 - 1	SU-03-041118	2	0.0157	-	7470A	PPB	13 Apr 2018 12:47:30	U	SMPL
J2335-02 - 1	S1	49	0.0288	-	7470A	PPB	13 Apr 2018 12:49:38	U	SMPL
J2335-04 - 1	S2	84	0.0386	-	7470A	PPB	13 Apr 2018 12:51:47	U	SMPL
J2335-06 - 1	S3	341	0.1103	-	7470A	PPB	13 Apr 2018 12:53:53	U	SMPL
CCV27 - 1	CCV27	18093	5.0616	-	7470A	PPB	13 Apr 2018 12:56:00	U	SMPL
CCB27 - 1	CCB27	-155	-0.0281	-	7470A	PPB	13 Apr 2018 12:58:08	U	SMPL
J2337-01 - 1	72-12002-SOIL	39	0.0260	-	7470A	PPB	13 Apr 2018 13:00:16	U	SMPL
J2337-03 - 1	72-12002-OILY-DEBRIS	-11	0.0121	-	7470A	PPB	13 Apr 2018 13:02:23	U	SMPL
J2337-05 - 1	ETGI-280-SOIL	3	0.0160	-	7470A	PPB	13 Apr 2018 13:04:30	U	SMPL
J2337-07 - 1	ETGI-280-DEBRIS	8	0.0174	-	7470A	PPB	13 Apr 2018 13:06:37	U	SMPL
J2337-09 - 1	RBR200036	19	0.0205	-	7470A	PPB	13 Apr 2018 13:08:44	U	SMPL
J2337-11 - 1	RBR200029	-5	0.0138	-	7470A	PPB	13 Apr 2018 13:10:52	U	SMPL
J2337-13 - 1	RBR200032	7	0.0171	-	7470A	PPB	13 Apr 2018 13:13:00	U	SMPL
J2337-15 - 1	VNJ-216-SPEEDY-DRY	6	0.0168	-	7470A	PPB	13 Apr 2018 13:15:08	U	SMPL
J2337-17 - 1	VNJ-216-DEBRIS	37	0.0255	-	7470A	PPB	13 Apr 2018 13:17:17	U	SMPL
PB108202BL - 1	PBW	3	0.0160	-	7470A	PPB	13 Apr 2018 13:19:24	U	SMPL
CCV28 - 1	CCV28	17755	4.9673	-	7470A	PPB	13 Apr 2018 13:21:31	U	SMPL
CCB28 - 1	CCB28	-155	-0.0281	-	7470A	PPB	13 Apr 2018 13:23:38	U	SMPL
PB108202BS - 1	LCSW	12108	3.3923	-	7470A	PPB	13 Apr 2018 13:25:46	U	SMPL
PB108202TB - 1	PB108168TB	-113	-0.0164	-	7470A	PPB	13 Apr 2018 13:27:53	U	SMPL
J2336-02 - 1	SL	22	0.0213	-	7470A	PPB	13 Apr 2018 13:30:00	U	SMPL
J2336-02DUP - 1	SLDUP	28	0.0230	-	7470A	PPB	13 Apr 2018 13:32:08	U	SMPL
J2336-02MS - 1	SLMS	13244	3.7091	-	7470A	PPB	13 Apr 2018 13:34:15	U	SMPL
J2336-02MSD - 1	SLMSD	12551	3.5158	-	7470A	PPB	13 Apr 2018 13:36:23	U	SMPL
PB108203BL - 1	PBW	-156	-0.0283	-	7470A	PPB	13 Apr 2018 13:38:31	U	SMPL
PB108203BS - 1	LCSW	12766	3.5758	-	7470A	PPB	13 Apr 2018 13:40:41	U	SMPL
J2314-01 - 1	CA-AQIDW01-20180409	-141	-0.0242	-	7470A	PPB	13 Apr 2018 13:42:49	U	SMPL
CCV29 - 1	CCV29	17649	4.9377	-	7470A	PPB	13 Apr 2018 13:47:03	U	SMPL
CCB29 - 1	CCB29	-163	-0.0303	-	7470A	PPB	13 Apr 2018 13:49:10	U	SMPL
J2336-01 - 1	SL	208	0.0732	-	7470A	PPB	13 Apr 2018 13:57:41	U	SMPL
CCV30 - 1	CCV30	17607	4.9260	-	7470A	PPB	13 Apr 2018 13:59:49	U	SMPL
CCB30 - 1	CCB30	-156	-0.0283	-	7470A	PPB	13 Apr 2018 14:01:57	U	SMPL
J2162-02LX5 - 1		0	0.0152	-	7470A	PPB	13 Apr 2018 14:04:06	U	SMPL
J2162-02A - 1		13175	3.6899	-	7470A	PPB	13 Apr 2018 14:06:14	U	SMPL
J2336-02LX5 - 1		-104	-0.0138	-	7470A	PPB	13 Apr 2018 14:08:22	U	SMPL
J2336-02A - 1		13102	3.6695	-	7470A	PPB	13 Apr 2018 14:10:30	U	SMPL
J2315-08 - 1	WW-DRUM-04092018	1074	0.3147	-	7470A	PPB	13 Apr 2018 14:15:11	U	SMPL
J2315-08DUP - 1	WW-DRUM-04092018DUP	1086	0.3181	-	7470A	PPB	13 Apr 2018 14:17:18	U	SMPL
J2315-08MS - 1	WW-DRUM-04092018MS	14294	4.0020	-	7470A	PPB	13 Apr 2018 14:19:26	U	SMPL
J2315-08MSD - 1	WW-DRUM-04092018MSD	13813	3.8678	-	7470A	PPB	13 Apr 2018 14:21:34	U	SMPL
J2315-08LX5 - 1		124	0.0498	-	7470A	PPB	13 Apr 2018 14:23:42	U	SMPL
J2315-08A - 1		13605	3.8098	-	7470A	PPB	13 Apr 2018 14:25:51	U	SMPL
CCV31 - 1	CCV31	17519	4.9015	-	7470A	PPB	13 Apr 2018 14:27:58	U	SMPL
CCB31 - 1	CCB31	-148	-0.0261	-	7470A	PPB	13 Apr 2018 14:30:06	U	SMPL