

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB94601BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP63677
calibration std. phosphate 0.5 ppm	WP63676
calibration std. phosphate 0.3 ppm	WP63675
calibration std. 0 ppm	WP63672
phosphate CCV std.	WP63678
calibration std. phosphate 0.1 ppm	WP63674
calibration std. phosphate 0.05 ppm	WP63673
Combined reagent	WP63683
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776
phosphate ICV std.	WP63679

Intercept: -0.001Slope: 0.6502Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/13/2018	11:46
2	CAL2	0.05	1	50	50	0.033	04/13/2018	11:46
3	CAL3	0.10	1	50	50	0.065	04/13/2018	11:47
4	CAL4	0.30	1	50	50	0.186	04/13/2018	11:47
5	CAL5	0.50	1	50	50	0.329	04/13/2018	11:48
6	CAL6	1.00	1	50	50	0.649	04/13/2018	11:48



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB94601

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	04/13/2018	11:49
2	ICB		1	50	50	0.000	0.002	04/13/2018	11:49
3	CCV1	0.50	1	50	50	0.327	0.504	04/13/2018	12:03
4	CCB1		1	50	50	0.000	0.002	04/13/2018	12:03
5	J2133-01		1	1.00	50	0.007	0.012	04/13/2018	12:05
6	J2133-02		1	1.00	50	0.036	0.057	04/13/2018	12:05
7	J2133-03		1	1.00	50	0.008	0.014	04/13/2018	12:06
8	CCV2	0.50	1	50	50	0.326	0.503	04/13/2018	12:07
9	CCB2		1	50	50	0.000	0.002	04/13/2018	12:07