

Daily Analysis Runlog For Sequence/QC Batch ID # LB94605

Review By		Review On				
SubDirectory		LB94605		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP63685,WP63686,WP63687,WP63688,WP63689,WP63690,WP63691				
ICV Standard		W1542				
CCV Standard		WP63693				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP62802				
Chk Standard		WP63692,WP59958,WP60878				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	04/13/18 12:37		
2	5.0PPBCN	5.0PPBCN	CAL2	04/13/18 12:37		
3	10PPBCN	10PPBCN	CAL3	04/13/18 12:37		
4	50PPBCN	50PPBCN	CAL4	04/13/18 12:37		
5	100PPBCN	100PPBCN	CAL5	04/13/18 12:37		
6	250PPBCN	250PPBCN	CAL6	04/13/18 12:37		
7	500PPBCN	500PPBCN	CAL7	04/13/18 12:37		
8	ICV1	ICV1	ICV	04/13/18 14:10		
9	ICB1	ICB1	ICB	04/13/18 14:10		
10	CCV1	CCV1	CCV	04/13/18 14:54		
11	CCB1	CCB1	CCB	04/13/18 14:54		
12	PB108227BL	PB108227BL	MB	04/13/18 14:54		
13	PB108227BS	PB108227BS	LCS	04/13/18 14:55		
14	LOWPB108227	LOWPB108227	SAM	04/13/18 14:55		
15	HIGHPB108227	HIGHPB108227	SAM	04/13/18 14:55		
16	J2152-01	BU-01-041218	SAM	04/13/18 14:55		
17	J2152-01DUP	BU-01-041218DUP	DUP	04/13/18 14:55		
18	J2152-01MS	BU-01-041218MS	MS	04/13/18 14:55		
19	J2152-01MSD	BU-01-041218MSD	MSD	04/13/18 14:55		
20	J2166-03	TR-04-041218	SAM	04/13/18 14:55		
21	J2348-01	135-348	SAM	04/13/18 15:02		

Daily Analysis Runlog For Sequence/QC Batch ID # LB94605

Review By		Review On				
SubDirectory		LB94605		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP63685,WP63686,WP63687,WP63688,WP63689,WP63690,WP63691				
ICV Standard		W1542				
CCV Standard		WP63693				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP62802				
Chk Standard		WP63692,WP59958,WP60878				
22	CCV2	CCV2	CCV	04/13/18 15:02		
23	CCB2	CCB2	CCB	04/13/18 15:02		
24	J2348-03	132	SAM	04/13/18 15:02		
25	J2348-05	155	SAM	04/13/18 15:02		
26	J2348-07	275	SAM	04/13/18 15:02		
27	J2348-09	281	SAM	04/13/18 15:02		
28	J2357-01	RBR2520258	SAM	04/13/18 15:02		
29	J2362-01	FK-BPM-01-D03-D03	SAM	04/13/18 15:02		
30	PB108228BL	PB108228BL	MB	04/13/18 15:02		
31	PB108228BS	PB108228BS	LCS	04/13/18 15:02		
32	J2364-01	SP-A	SAM	04/13/18 15:09		
33	J2364-01DUP	SP-ADUP	DUP	04/13/18 15:09		
34	CCV3	CCV3	CCV	04/13/18 15:09		
35	CCB3	CCB3	CCB	04/13/18 15:09		
36	J2364-01MS	SP-AMS	MS	04/13/18 15:09		
37	J2364-01MSD	SP-AMSD	MSD	04/13/18 15:09		
38	J2364-05	SP-B	SAM	04/13/18 15:09		
39	J2364-09	SP-C	SAM	04/13/18 15:09		
40	CCV4	CCV4	CCV	04/13/18 15:09		
41	CCB4	CCB4	CCB	04/13/18 15:09		



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB94605	Date Received:	
Client Sample ID:	HIGHPB108227	SDG No.:	LB94605
Lab Sample ID:	HIGHPB108227	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	BU-01-041218	SDG No.:	LB94605
Lab Sample ID:	J2152-01	Matrix:	Solid
Level (low/med):	low	% Solid:	88.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.1	J	1	0.036		0.28	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	TR-04-041218	SDG No.:	LB94605
Lab Sample ID:	J2166-03	Matrix:	Solid
Level (low/med):	low	% Solid:	87.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.038	J	1	0.036		0.27	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	135-348	SDG No.:	LB94605
Lab Sample ID:	J2348-01	Matrix:	Solid
Level (low/med):	low	% Solid:	87.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.029	U	1	0.037		0.28	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	132	SDG No.:	LB94605
Lab Sample ID:	J2348-03	Matrix:	Solid
Level (low/med):	low	% Solid:	88.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.029	U	1	0.036		0.27	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	155	SDG No.:	LB94605
Lab Sample ID:	J2348-05	Matrix:	Solid
Level (low/med):	low	% Solid:	88.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.072	J	1	0.036		0.28	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	275	SDG No.:	LB94605
Lab Sample ID:	J2348-07	Matrix:	Solid
Level (low/med):	low	% Solid:	89

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.053	J	1	0.036		0.27	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	281	SDG No.:	LB94605
Lab Sample ID:	J2348-09	Matrix:	Solid
Level (low/med):	low	% Solid:	87.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.29		1	0.035		0.27	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	RBR2520258	SDG No.:	LB94605
Lab Sample ID:	J2357-01	Matrix:	Solid
Level (low/med):	low	% Solid:	88.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.055	J	1	0.036		0.27	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	FK-BPM-01-D03-D03A-E03-E03A	SDG No.:	LB94605
Lab Sample ID:	J2362-01	Matrix:	Solid
Level (low/med):	low	% Solid:	89.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.02	U	1	0.034		0.26	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	SP-A	SDG No.:	LB94605
Lab Sample ID:	J2364-01	Matrix:	Solid
Level (low/med):	low	% Solid:	88.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.018	U	1	0.036		0.27	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	SP-B	SDG No.:	LB94605
Lab Sample ID:	J2364-05	Matrix:	Solid
Level (low/med):	low	% Solid:	90.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.062	J	1	0.033		0.25	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	4/12/2018 12:00:00 AM
Project:	LB94605	Date Received:	4/12/2018 12:00:00 AM
Client Sample ID:	SP-C	SDG No.:	LB94605
Lab Sample ID:	J2364-09	Matrix:	Solid
Level (low/med):	low	% Solid:	94.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.12	J	1	0.032		0.24	mg/Kg	04/13/2018	04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB94605	Date Received:	
Client Sample ID:	LOWPB108227	SDG No.:	LB94605
Lab Sample ID:	LOWPB108227	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.011		1	0.003		0.005	mg/L		04/13/2018	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB94605	Date Received:	
Client Sample ID:	LOWPB108227	SDG No.:	LB94605
Lab Sample ID:	LOWPB108227	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94605

SDG No.: LB94605

Contract: _____

Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Cyanide	0.098	0.099	99		85 - 115		04/13/2018	14:10	lb94605



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB94605

SDG No.: LB94605

Contract: _____ Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.23	0.25	92		90 - 110		04/13/2018	14:54	lb94605
CCV2	Cyanide	0.24	0.25	96		90 - 110		04/13/2018	15:02	lb94605
CCV3	Cyanide	0.24	0.25	96		90 - 110		04/13/2018	15:09	lb94605
CCV4	Cyanide	0.24	0.25	96		90 - 110		04/13/2018	15:09	lb94605



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94605

SDG No.: LB94605

Contract:

Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	0.00024	+/-0.005	U		0.005		04/13/2018	14:10	lb94605



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94605

SDG No.: LB94605

Contract:

Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	0.00041	+/-0.005	U		0.005		04/13/2018	14:54	lb94605
CCB2	Cyanide	0.00071	+/-0.005	U		0.005		04/13/2018	15:02	lb94605
CCB3	Cyanide	0.00069	+/-0.005	U		0.005		04/13/2018	15:09	lb94605
CCB4	Cyanide	0.00027	+/-0.005	U		0.005		04/13/2018	15:09	lb94605



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB94605

SDG No.: LB94605

Contract: Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB94605

SDG No.: LB94605

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB108227BL	Cyanide	SOLID		Batch Number:		PB108227		Prep Date:	04/13/2018	
		-0.013	<0.25	U		0.25		04/13/2018	14:54	lb94605
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB108228BL	Cyanide	SOLID		Batch Number:		PB108228		Prep Date:	04/13/2018	
		0.0013	<0.25	U		0.25		04/13/2018	15:02	lb94605



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB94605 level: low sdg no.: LB94605
contract: lab code: CHEM case no.: LB94605 sas no.: LB94605
matrix: Solid sample id: J2152-01 client id: BU-01-041218MS
Percent Solids for Sample: 88.9 Spiked ID: J2152-01MS Percent Solids for Spike Sample: 88.9

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.5		0.1	J	2.2	64	*	



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB94605 level: low sdg no.: LB94605
contract: lab code: CHEM case no.: LB94605 sas no.: LB94605
matrix: Solid sample id: J2152-01 client id: BU-01-041218MSD
Percent Solids for Sample: 88.9 Spiked ID: J2152-01MSD Percent Solids for Spike Sample: 88.9

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.5		0.1	J	2.2	64	*	



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB94605 level: low sdg no.: LB94605
contract: lab code: CHEM case no.: LB94605 sas no.: LB94605
matrix: Solid sample id: J2364-01 client id: SP-AMS
Percent Solids for Sample: 88.4 Spiked ID: J2364-01MS Percent Solids for Spike Sample: 88.4

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	2		0.036	U	2.2	90.9		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB94605 level: low sdg no.: LB94605
contract: lab code: CHEM case no.: LB94605 sas no.: LB94605
matrix: Solid sample id: J2364-01 client id: SP-AMSD
Percent Solids for Sample: 88.4 Spiked ID: J2364-01MSD Percent Solids for Spike Sample: 88.4

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125		2	0.036	U	2.2	90.9		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB94605 **Level:** LOW **SDG No.:** LB94605
Contract: **Lab Code:** CHEM **Case No.:** LB94605 **SAS No.:** LB94605
Matrix: Solid **Sample ID:** J2152-01 **Client ID:** BU-01-041218DUP
Percent Solids for Sample: 88.9 **Duplicate ID** J2152-01DUP **Percent Solids for Spike Sample:** 88.9

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.1	J	0.063	J	45	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB94605 **Level:** LOW **SDG No.:** LB94605
Contract: **Lab Code:** CHEM **Case No.:** LB94605 **SAS No.:** LB94605
Matrix: Solid **Sample ID:** J2152-01MS **Client ID:** BU-01-041218MSD
Percent Solids for Sample: 88.9 **Duplicate ID** J2152-01MSD **Percent Solids for Spike Sample:** 88.9

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.5		1.5		0		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB94605 **Level:** LOW **SDG No.:** LB94605
Contract: **Lab Code:** CHEM **Case No.:** LB94605 **SAS No.:** LB94605
Matrix: Solid **Sample ID:** J2364-01 **Client ID:** SP-ADUP
Percent Solids for Sample: 88.4 **Duplicate ID** J2364-01DUP **Percent Solids for Spike Sample:** 88.4

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.036	U	0.062	J	200.0		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB94605 **Level:** LOW **SDG No.:** LB94605
Contract: **Lab Code:** CHEM **Case No.:** LB94605 **SAS No.:** LB94605
Matrix: Solid **Sample ID:** J2364-01MS **Client ID:** SP-AMSD
Percent Solids for Sample: 88.4 **Duplicate ID** J2364-01MSD **Percent Solids for Spike Sample:** 88.4

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	2		2		0		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB94605

SDG No.: LB94605

Contract:

Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB108227BS Cyanide	mg/Kg	4.9	4.9		100	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB94605

SDG No.: LB94605

Contract:

Lab Code: CHEM

Case No.: LB94605

SAS No.: LB94605

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB108228BS Cyanide	mg/Kg	4.9	4.9		100	85 - 115	