


7471B

LB 94608
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42349



A= 0.0000e+000

B= 2.3625e-004

C= -2.2190e-002

Rho= 0.9999733

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	19	0.000	19				
0.2	0.200	0.190	-0.010	897	0.0 %	897				
2.5	2.500	2.505	0.005	10697	0.0 %	10697				
5.0	5.000	5.058	0.058	21502	0.0 %	21502				
7.5	7.500	7.483	-0.017	31766	0.0 %	31766				
10	10.000	9.983	-0.017	42349	0.0 %	42349				

LB94608 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Apr 2018 17:23:48

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	19	-	0.00007471B	PPB	13 Apr 2018 17:25:40	S	Std
0.2 - 1	50.2	897	-	0.20007471B	PPB	13 Apr 2018 17:27:58	S	Std
2.5 - 1	52.5	10697	-	2.50007471B	PPB	13 Apr 2018 17:30:15	S	Std
5.0 - 1	55	21502	-	5.00007471B	PPB	13 Apr 2018 17:32:33	S	Std
7.5 - 1	57.5	31766	-	7.50007471B	PPB	13 Apr 2018 17:34:50	S	Std
10 - 1	510	42349	-	10.00007471B	PPB	13 Apr 2018 17:39:40	S	Std
ICV94 - 1	ICV94	16428	3.8589	-7471B	PPB	13 Apr 2018 17:42:21	U	SMPL
ICB94 - 1	ICB94	-140	-0.0553	-7471B	PPB	13 Apr 2018 17:44:37	U	SMPL
CCV36 - 1	CCV36	22126	5.2051	-7471B	PPB	13 Apr 2018 17:46:54	U	SMPL
CCB36 - 1	CCB36	-165	-0.0612	-7471B	PPB	13 Apr 2018 17:49:10	U	SMPL
CRA - 1	CRA	850	0.1786	-7471B	PPB	13 Apr 2018 17:51:26	U	SMPL
HighStd - 1	HighStd	42535	10.0267	-7471B	PPB	13 Apr 2018 17:53:42	U	SMPL
ChkStd - 1	ChkStd	29008	6.8310	-7471B	PPB	13 Apr 2018 17:55:58	U	SMPL
PB108212BL - 1	PBS	-36	-0.0307	-7471B	PPB	13 Apr 2018 18:01:03	U	SMPL
PB108212BS - 1	LCSS	14565	3.4188	-7471B	PPB	13 Apr 2018 18:03:19	U	SMPL
J2152-01 - 1	BU-01-041218	2612	0.5949	-7471B	PPB	13 Apr 2018 18:05:36	U	SMPL
J2166-03 - 1	TR-04-041218	1879	0.4217	-7471B	PPB	13 Apr 2018 18:07:53	U	SMPL
J2348-01 - 1	135-348	3184	0.7300	-7471B	PPB	13 Apr 2018 18:10:11	U	SMPL
J2348-03 - 1	132	4020	0.9275	-7471B	PPB	13 Apr 2018 18:12:28	U	SMPL
J2348-05 - 1	155	7668	1.7894	-7471B	PPB	13 Apr 2018 18:14:44	U	SMPL
CCV37 - 1	CCV37	21734	5.1125	-7471B	PPB	13 Apr 2018 18:17:00	U	SMPL
CCB37 - 1	CCB37	-120	-0.0505	-7471B	PPB	13 Apr 2018 18:19:16	U	SMPL
J2348-07 - 1	275	8021	1.8728	-7471B	PPB	13 Apr 2018 18:21:33	U	SMPL
J2348-09 - 1	281	4113	0.9495	-7471B	PPB	13 Apr 2018 18:23:49	U	SMPL
J2357-01 - 1	RBR2520258	55421	13.0710	-7471B	PPB	13 Apr 2018 18:26:05	U	SMPL
J2360-01 - 1	NAA-WS-001	55	-0.0092	-7471B	PPB	13 Apr 2018 18:28:21	U	SMPL
J2360-02 - 1	NAA-WS-002	301	0.0489	-7471B	PPB	13 Apr 2018 18:30:41	U	SMPL
J2360-03 - 1	NAA-WS-003	354	0.0614	-7471B	PPB	13 Apr 2018 18:32:58	U	SMPL
J2360-04 - 1	NAA-WS-004	387	0.0692	-7471B	PPB	13 Apr 2018 18:35:15	U	SMPL
J2362-01 - 1	FK-BPM-01-D03-D03A-E03-E03A	1195	0.2601	-7471B	PPB	13 Apr 2018 18:37:33	U	SMPL
J2362-01DUP - 1	FK-BPM-01-D03-D03A-E03-E03ADUP	1368	0.3010	-7471B	PPB	13 Apr 2018 18:39:51	U	SMPL
J2362-01MS - 1	FK-BPM-01-D03-D03A-E03-E03AMS	17289	4.0623	-7471B	PPB	13 Apr 2018 18:42:07	U	SMPL
CCV38 - 1	CCV38	22002	5.1758	-7471B	PPB	13 Apr 2018 18:44:24	U	SMPL
CCB38 - 1	CCB38	-138	-0.0548	-7471B	PPB	13 Apr 2018 18:46:40	U	SMPL
J2362-01MSD - 1	FK-BPM-01-D03-D03A-E03-E03AMSD	17453	4.1011	-7471B	PPB	13 Apr 2018 18:48:57	U	SMPL
J2363-01 - 1	EME-4-12	1116	0.2415	-7471B	PPB	13 Apr 2018 18:51:14	U	SMPL
J2364-01 - 1	SP-A	3397	0.7804	-7471B	PPB	13 Apr 2018 18:53:30	U	SMPL
J2364-05 - 1	SP-B	5515	1.2807	-7471B	PPB	13 Apr 2018 18:55:47	U	SMPL
J2364-09 - 1	SP-C	1943	0.4368	-7471B	PPB	13 Apr 2018 18:58:04	U	SMPL
J2364-13 - 1	CO1	201	0.0253	-7471B	PPB	13 Apr 2018 19:00:21	U	SMPL
J2364-15 - 1	CO2	2064	0.4654	-7471B	PPB	13 Apr 2018 19:02:38	U	SMPL
J2362-01LX5 - 1		238	0.0340	-7471B	PPB	13 Apr 2018 19:04:56	U	SMPL
J2362-01A - 1		17516	4.1160	-7471B	PPB	13 Apr 2018 19:07:14	U	SMPL
J2357-01DLX2 - 1		27168	6.3963	-7471B	PPB	13 Apr 2018 19:09:30	U	SMPL
CCV39 - 1	CCV39	21839	5.1373	-7471B	PPB	13 Apr 2018 19:11:46	U	SMPL
CCB39 - 1	CCB39	-90	-0.0435	-7471B	PPB	13 Apr 2018 19:14:03	U	SMPL