



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94613

Slope : 97.2

BalanceID: WC SC-4

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.3	4.01	04/13/2018	16:30
2	CAL2	1	Water			20.4	7.00	04/13/2018	16:32
3	CAL3	1	Water			20.4	10.01	04/13/2018	16:33
4	ICV	1	Water			20.4	7.01	04/13/2018	16:35
5	CCV1	1	Water			20.2	2.01	04/13/2018	16:36
6	J2369-04	1	Solid	20.02	20	22.5	5.89	04/13/2018	16:38
7	J2369-04DUP	1	Solid	20.03	20	22.4	5.90	04/13/2018	16:40
8	J2370-08	1	Solid	20.04	20	22.6	9.27	04/13/2018	16:42
9	CCV2	1	Water			22.1	12.01	04/13/2018	16:44

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j12386

WorkList ID : 110773

Date : 4/13/2018 3:57:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/13/2018	Solid	J2369-04	Corrosivity	Cool 4 deg C	PSEG01	D11	RO-1-4	04/13/2018	9045C
04/13/2018	Solid	J2370-08	Corrosivity	Cool 4 deg C	PSEG01	D11	CG	04/13/2018	9045C

Date/Time 04-13-18 4:00PM
 Received by: 30
 Relinquished by: SG

Date/Time 04-13-18 5:20PM
 Received by: CP
 Relinquished by: JP