



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94625

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	04/13/2018	16:25
2	CAL2	1	Water			22.3	7.01	04/13/2018	16:27
3	CAL3	1	Water			22.0	10.02	04/13/2018	16:29
4	ICV	1	Water			22.3	7.01	04/13/2018	16:31
5	CCV1	1	Water			22.5	2.00	04/13/2018	16:33
6	J2373-01	1	Water			20.6	3.51	04/13/2018	16:35
7	J2373-01DUP	1	Water			20.6	3.52	04/13/2018	16:37
8	J2383-04	1	Water			21.0	7.41	04/13/2018	16:39
9	CCV2	1	Water			22.3	12.00	04/13/2018	16:42

WORKLIST(Hardcopy Internal Chain)

LB94625

Worklist Name : pH J2373

Worklist ID : 110770

Date : 4/13/2018 2:56:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/12/2018	Water	J2373-01	pH	Cool 4 deg C	STER01	N13	POOL-WATER	04/12/2018	SM 4500-PH B
04/12/2018	Water	J2383-04	pH	Cool 4 deg C	THEP04	N13	DRUM-PW-COMPOSITE	04/12/2018	SM 4500-PH B

Date/Time

04/13/18 4:40 PM

Received by:

SP

Relinquished by:

CP

Date/Time

04/13/18 4:55 PM

Received by:

CP

Relinquished by:

SP