



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94626

Ambient Barometric Pressure (mmHg): 750

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	83.9	04/16/2018	08:45
2	J2383-04		1	>100.0	>212.0	>212.0	04/16/2018	09:20
3	J2383-04DUP		1	>100.0	>212.0	>212.0	04/16/2018	09:45
4	J2390-01		1	>100.0	>212.0	>212.0	04/16/2018	10:22

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB94626

WorkList Name : FP-041618

WorkList ID : 110802

Date : 4/16/2018 8:10:44 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/20/2018	Water	J2383-04	Flash Point	Cool 4 deg C	THEP04	N13	DRUM-PW-COMPOSITE	04/12/2018	1010A
04/27/2018	Water	J2390-01	Flash Point	Cool 4 deg C	DIST01	N13	2AVE-12	04/13/2018	1010A

Date/Time 04/16/18 8:15 AM

Received by: KS

Relinquished by: CP

Date/Time 04/16/18 10:25 AM

Received by: CP

Relinquished by: KS