



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 04/17/2018OVEN TEMP1: 103 °C TEMP: 104 °CDATE & TIME IN: 04/16/2018 15:50 OUT: 04/17/2018 08:40OVEN TEMP2: 105 °C TEMP: 105 °CDATE & TIME IN: 04/17/2018 09:20 OUT: 04/17/2018 10:40Run Number: LB94645BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB94645BL	LB94645BL	81.4942	81.4942	100	81.4942	81.4942	81.4942	0.0000	0
2	J2390-01	2AVE-12	63.8370	63.8370	100	63.9690	63.9690	63.9690	0.1320	1320
3	J2390-01DUP	2AVE-12DUP	84.9878	84.9878	100	85.1250	85.1250	85.1250	0.1372	1372

A = Sample Volume (ml)
B = Empty Dish Weight (g)
C = Final Empty Dish + Sample weight after drying @105°C(g)
D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS041618

WorkList ID : 110847

Date : 4/16/2018 2:31:30 PM

21394645

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/20/2018	Water	J2390-01	TS	Cool 4 deg C	DIST01	N13	2AVE-12	04/13/2018	SM2540 B

Date/Time 04/16/18, 15:00
 Received by: AS
 Relinquished by: AS

Date/Time 04/16/18, 16:00
 Received by: AS
 Relinquished by: AS