

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/16/2018 15:18

Test: Reactive ~~Total~~ CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.174	0.0	0.097	
ICB1	0.399	0.0	0.004	
CCV1	247.097	0.0	0.243	
CCB1	0.412	0.0	0.004	
PB108252BL	0.261	0.0	0.004	
J2369-04	0.280	0.0	0.004	
J2369-04DUP	0.155	0.0	0.004	
J2370-08	0.446	0.0	0.004	
CCV2	248.649	0.0	0.244	
CCB2	0.165	0.0	0.004	

N	10
Mean	59.404
SD	103.7336
CV%	174.62

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 16 15:11:21 2018

Mon Apr 16 15:11:30 2018

AS

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.0917	µg/l	4/16/2018 12:33:44	
5.OPPBCN	A	Reactive CN	P	4.4834	µg/l	4/16/2018 12:33:45	
10PPBCN	A	Reactive CN	P	9.8327	µg/l	4/16/2018 12:33:46	
50PPBCN	A	Reactive CN	P	50.9739	µg/l	4/16/2018 12:33:47	
100PPBCN	A	Reactive CN	P	101.3767	µg/l	4/16/2018 12:33:48	
250PPBCN	A	Reactive CN	P	247.5786	µg/l	4/16/2018 12:33:49	
500PPBCN	A	Reactive CN	P	500.8465	µg/l	4/16/2018 12:33:50	
ICV1	S	Reactive CN	P	96.1741	µg/l	4/16/2018 15:11:21	
ICB1	S	Reactive CN	P	0.3987	µg/l	4/16/2018 15:11:22	
CCV1	S	Reactive CN	P	247.0974	µg/l	4/16/2018 15:11:23	
CCB1	S	Reactive CN	P	0.4116	µg/l	4/16/2018 15:11:24	
PB108252BL	S	Reactive CN	P	0.2606	µg/l	4/16/2018 15:11:25	
J2369-04	S	Reactive CN	P	0.2796	µg/l	4/16/2018 15:11:26	
J2369-04DUP	S	Reactive CN	P	0.1548	µg/l	4/16/2018 15:11:27	
J2370-08	S	Reactive CN	P	0.4456	µg/l	4/16/2018 15:11:28	
CCV2	S	Reactive CN	P	248.6493	µg/l	4/16/2018 15:11:29	
CCB2	S	Reactive CN	P	0.1645	µg/l	4/16/2018 15:11:30	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

4/16/2018 12:36


Test Total CN

Accepted 4/16/2018 12:36

Factor 1033
 Bias 0.003

Coeff. of det. 0.999953

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	0.0PPBCN	0.003	-0.0917	0.0000	
2	5.0PPBCN	0.008	4.4834	5.0000	
3	10PPBCN	0.013	9.8327	10.0000	
4	50PPBCN	0.053	50.9739	50.0000	
5	100PPBCN	0.102	101.3767	100.0000	
6	250PPBCN	0.243	247.5786	250.0000	
7	500PPBCN	0.488	500.8465	500.0000	