



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94652

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	04/16/2018	16:15
2	CAL2	1	Water			22.1	7.01	04/16/2018	16:17
3	CAL3	1	Water			22.3	10.02	04/16/2018	16:19
4	ICV	1	Water			22.5	7.01	04/16/2018	16:21
5	CCV1	1	Water			22.1	2.00	04/16/2018	16:23
6	J2410-04	1	Solid	20.2	20	20.6	11.43	04/16/2018	16:25
7	J2410-04DUP	1	Solid	20.5	20	20.6	11.44	04/16/2018	16:27
8	J2410-08	1	Solid	20.0	20	20.6	10.85	04/16/2018	16:29
9	J2411-03	1	Solid	20.5	20	20.8	9.10	04/16/2018	16:31
10	CCV2	1	Water			22.5	12.00	04/16/2018	16:33

WORKLIST(Hardcopy Internal Chain)

12B 04652

Worklist Name : corrosivity j2410

Worklist ID : 110853

Date : 4/16/2018 4:02:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/16/2018	Solid	J2410-04	Corrosivity	Cool 4 deg C	PSEG01	D21	T-2	04/16/2018	9045C
04/16/2018	Solid	J2410-08	Corrosivity	Cool 4 deg C	PSEG01	D21	T-3	04/16/2018	9045C
04/16/2018	Solid	J2411-03	Corrosivity	Cool 4 deg C	PSEG01	D22	DRUM-1-3	04/16/2018	9045C

Date/Time 04-16-18 4:00 PM
 Received by: JS
 Relinquished by: CP

Date/Time 04-16-18 5:25 PM
 Received by: CP
 Relinquished by: JS