


245.1

LA94664
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34048



A= 0.0000e+000

B= 2.9676e-004

C= -4.6189e-002

Rho= 0.9998033

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	77	0.000	77				
0.05	0.050	0.050	0.000	325	0.0 %	325				
0.2	0.200	0.161	-0.039	697	0.0 %	697				
2.5	2.500	2.618	0.118	8976	0.0 %	8976				
5.0	5.000	5.022	0.022	17079	0.0 %	17079				
7.5	7.500	7.365	-0.135	24974	0.0 %	24974				
10.0	10.000	10.058	0.058	34048	0.0 %	34048				

LB94664 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Apr 2018 10:40:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	77	-	0.0000	245.1	PPB	17 Apr 2018 11:03:27	S	Std
0.05 - 1	5005	325	-	0.0500	245.1	PPB	17 Apr 2018 11:05:35	S	Std
0.2 - 1	502	697	-	0.2000	245.1	PPB	17 Apr 2018 11:07:44	S	Std
2.5 - 1	52.5	8976	-	2.5000	245.1	PPB	17 Apr 2018 11:09:52	S	Std
5.0 - 1	55	17079	-	5.0000	245.1	PPB	17 Apr 2018 11:12:00	S	Std
7.5 - 1	57.5	24974	-	7.5000	245.1	PPB	17 Apr 2018 11:16:37	S	Std
10.0 - 1	510	34048	-	10.0000	245.1	PPB	17 Apr 2018 11:21:09	S	Std
ICV97 - 1	ICV97	13440	3.9422	-245.1		PPB	17 Apr 2018 11:31:11	U	SMPL
ICB97 - 1	ICB97	58	-0.0290	-245.1		PPB	17 Apr 2018 11:33:17	U	SMPL
CCV49 - 1	CCV49	16831	4.9485	-245.1		PPB	17 Apr 2018 11:35:24	U	SMPL
CCB49 - 1	CCB49	9	-0.0435	-245.1		PPB	17 Apr 2018 11:37:31	U	SMPL
CRA - 1	CRA	312	0.0464	-245.1		PPB	17 Apr 2018 11:42:06	U	SMPL
HighStd - 1	HighStd	33304	9.8370	-245.1		PPB	17 Apr 2018 11:44:12	U	SMPL
ChkStd - 1	ChkStd	22975	6.7718	-245.1		PPB	17 Apr 2018 11:46:19	U	SMPL
PB108263BL - 1	PBW	87	-0.0204	-245.1		PPB	17 Apr 2018 11:51:36	U	SMPL
PB108263BS - 1	LCSW	12320	3.6099	-245.1		PPB	17 Apr 2018 11:53:43	U	SMPL
J2391-01 - 1	041218-DBRI-F-D-B	-87	-0.0720	-245.1		PPB	17 Apr 2018 11:55:50	U	SMPL
J2391-02 - 1	041218-DBRI-F-D-M	4	-0.0450	-245.1		PPB	17 Apr 2018 11:57:58	U	SMPL
J2391-03 - 1	041218-DBRI-F-D-S	1	-0.0459	-245.1		PPB	17 Apr 2018 12:00:07	U	SMPL
J2391-04 - 1	041218-DBRI-F-U-B	37	-0.0352	-245.1		PPB	17 Apr 2018 12:02:15	U	SMPL
J2391-05 - 1	041218-DBRI-F-U-M	12	-0.0426	-245.1		PPB	17 Apr 2018 12:04:22	U	SMPL
CCV50 - 1	CCV50	16858	4.9565	-245.1		PPB	17 Apr 2018 12:06:29	U	SMPL
CCB50 - 1	CCB50	11	-0.0429	-245.1		PPB	17 Apr 2018 12:08:35	U	SMPL
J2391-06 - 1	041218-DBRI-F-U-S	-19	-0.0518	-245.1		PPB	17 Apr 2018 12:10:43	U	SMPL
J2391-07 - 1	041218-JETB-E-D-B	25	-0.0388	-245.1		PPB	17 Apr 2018 12:12:50	U	SMPL
J2391-08 - 1	041218-JETB-E-D-M	15	-0.0417	-245.1		PPB	17 Apr 2018 12:14:57	U	SMPL
J2391-09 - 1	041218-JETB-E-D-S	12	-0.0426	-245.1		PPB	17 Apr 2018 12:17:04	U	SMPL
J2391-10 - 1	041218-JETB-E-U-B	15	-0.0417	-245.1		PPB	17 Apr 2018 12:19:12	U	SMPL
J2391-11 - 1	041218-JETB-E-U-M	0	-0.0462	-245.1		PPB	17 Apr 2018 12:21:19	U	SMPL
J2391-12 - 1	041218-JETB-E-U-S	10	-0.0432	-245.1		PPB	17 Apr 2018 12:23:28	U	SMPL
J2391-13 - 1	041218-SIDI-E-D-M	0	-0.0462	-245.1		PPB	17 Apr 2018 12:25:36	U	SMPL
J2391-14 - 1	041218-SIDI-E-U-M	26	-0.0385	-245.1		PPB	17 Apr 2018 12:27:44	U	SMPL
J2391-14DUP - 1	041218-SIDI-E-U-MDUP	17	-0.0411	-245.1		PPB	17 Apr 2018 12:29:51	U	SMPL
CCV51 - 1	CCV51	16743	4.9224	-245.1		PPB	17 Apr 2018 12:31:58	U	SMPL
CCB51 - 1	CCB51	18	-0.0408	-245.1		PPB	17 Apr 2018 12:34:05	U	SMPL
J2391-14MS - 1	041218-SIDI-E-U-MMS	14239	4.1793	-245.1		PPB	17 Apr 2018 12:36:14	U	SMPL
J2391-14MSD - 1	041218-SIDI-E-U-MMSD	14982	4.3998	-245.1		PPB	17 Apr 2018 12:38:21	U	SMPL
PB108264BL - 1	PBW	-143	-0.0886	-245.1		PPB	17 Apr 2018 12:40:28	U	SMPL
PB108264BS - 1	LCSW	13928	4.0870	-245.1		PPB	17 Apr 2018 12:42:35	U	SMPL
J2390-01 - 1	2AVE-12	23	-0.0394	-245.1		PPB	17 Apr 2018 12:44:42	U	SMPL
J2390-01DUP - 1	2AVE-12DUP	66	-0.0266	-245.1		PPB	17 Apr 2018 12:46:50	U	SMPL
J2390-01MS - 1	2AVE-12MS	14975	4.3977	-245.1		PPB	17 Apr 2018 12:48:58	U	SMPL
J2390-01MSD - 1	2AVE-12MSD	14605	4.2879	-245.1		PPB	17 Apr 2018 12:51:06	U	SMPL
J2392-01 - 1	041318-JETB-F-D-B	-72	-0.0676	-245.1		PPB	17 Apr 2018 12:53:15	U	SMPL
J2392-02 - 1	041318-JETB-F-D-M	8	-0.0438	-245.1		PPB	17 Apr 2018 12:55:22	U	SMPL
CCV52 - 1	CCV52	16464	4.8396	-245.1		PPB	17 Apr 2018 12:57:29	U	SMPL
CCB52 - 1	CCB52	16	-0.0414	-245.1		PPB	17 Apr 2018 12:59:37	U	SMPL
J2392-03 - 1	041318-JETB-F-D-S	-21	-0.0524	-245.1		PPB	17 Apr 2018 13:01:45	U	SMPL
J2392-04 - 1	041318-JETB-F-U-B	24	-0.0391	-245.1		PPB	17 Apr 2018 13:03:52	U	SMPL
J2392-05 - 1	041318-JETB-F-U-M	6	-0.0444	-245.1		PPB	17 Apr 2018 13:05:59	U	SMPL
J2392-06 - 1	041318-JETB-F-U-S	16	-0.0414	-245.1		PPB	17 Apr 2018 13:08:07	U	SMPL
J2397-02 - 1	EFF-WW	372	0.0642	-245.1		PPB	17 Apr 2018 13:10:14	U	SMPL
J2391-14LX5 - 1		122	-0.0100	-245.1		PPB	17 Apr 2018 13:12:22	U	SMPL
J2391-14A - 1		14597	4.2856	-245.1		PPB	17 Apr 2018 13:14:29	U	SMPL
J2390-01LX5 - 1		56	-0.0296	-245.1		PPB	17 Apr 2018 13:16:38	U	SMPL
J2390-01A - 1		14936	4.3862	-245.1		PPB	17 Apr 2018 13:18:47	U	SMPL
CCV53 - 1	CCV53	16692	4.9073	-245.1		PPB	17 Apr 2018 13:20:54	U	SMPL
CCB53 - 1	CCB53	7	-0.0441	-245.1		PPB	17 Apr 2018 13:23:01	U	SMPL