


7470A

LB94676
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

37106



A= 0.0000e+000

B= 2.7278e-004

C= 2.5333e-002

Rho= 0.9996215

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	21	0.000	21				
0.2	0.200	0.220	0.020	712	0.0 %	712				
2.5	2.500	2.527	0.027	9171	0.0 %	9171				
5.0	5.000	4.968	-0.032	18118	0.0 %	18118				
7.5	7.500	7.308	-0.192	26696	0.0 %	26696				
10.0	10.000	10.147	0.147	37106	0.0 %	37106				

LB94676 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Apr 2018 14:04:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	21	-	0.0000	7470A	PPB	17 Apr 2018 14:07:07	S	Std
0.2 - 1	SO.2	712	-	0.2000	7470A	PPB	17 Apr 2018 14:09:15	S	Std
2.5 - 1	SO.5	9171	-	2.5000	7470A	PPB	17 Apr 2018 14:11:24	S	Std
5.0 - 1	SO	18118	-	5.0000	7470A	PPB	17 Apr 2018 14:13:32	S	Std
7.5 - 1	SO.5	26696	-	7.5000	7470A	PPB	17 Apr 2018 14:15:40	S	Std
10.0 - 1	SO	37106	-	10.0000	7470A	PPB	17 Apr 2018 14:17:47	S	Std
ICV98 - 1	ICV98	14149	3.8849	-7470A	PPB	PPB	17 Apr 2018 14:20:21	U	SMPL
ICB98 - 1	ICB98	-159	-0.0180	-7470A	PPB	PPB	17 Apr 2018 14:22:27	U	SMPL
CCV54 - 1	CCV54	18173	4.9826	-7470A	PPB	PPB	17 Apr 2018 14:24:34	U	SMPL
CCB54 - 1	CCB54	-161	-0.0186	-7470A	PPB	PPB	17 Apr 2018 14:26:41	U	SMPL
CRA - 1	CRA	723	0.2226	-7470A	PPB	PPB	17 Apr 2018 14:28:48	U	SMPL
HighStd - 1	HighStd	36978	10.1123	-7470A	PPB	PPB	17 Apr 2018 14:30:55	U	SMPL
ChkStd - 1	ChkStd	25962	7.1073	-7470A	PPB	PPB	17 Apr 2018 14:33:02	U	SMPL
PB108269BL - 1	PBW	-206	-0.0309	-7470A	PPB	PPB	17 Apr 2018 14:35:11	U	SMPL
PB108269BS - 1	LCSW	13895	3.8156	-7470A	PPB	PPB	17 Apr 2018 14:37:20	U	SMPL
PB108269TB - 1	PB108243TB	-79	0.0038	-7470A	PPB	PPB	17 Apr 2018 14:39:28	U	SMPL
J2369-04 - 1	RO-1-4	37	0.0354	-7470A	PPB	PPB	17 Apr 2018 14:41:36	U	SMPL
J2369-04DUP - 1	RO-1-4DUP	45	0.0376	-7470A	PPB	PPB	17 Apr 2018 14:43:44	U	SMPL
J2369-04MS - 1	RO-1-4MS	13203	3.6269	-7470A	PPB	PPB	17 Apr 2018 14:45:52	U	SMPL
J2369-04MSD - 1	RO-1-4MSD	13701	3.7627	-7470A	PPB	PPB	17 Apr 2018 14:47:59	U	SMPL
CCV55 - 1	CCV55	18054	4.9502	-7470A	PPB	PPB	17 Apr 2018 14:50:06	U	SMPL
CCB55 - 1	CCB55	-173	-0.0219	-7470A	PPB	PPB	17 Apr 2018 14:52:13	U	SMPL
J2370-02 - 1	CG-RO-721194	57	0.0409	-7470A	PPB	PPB	17 Apr 2018 14:54:22	U	SMPL
J2370-04 - 1	CG-RO-7211981	39	0.0360	-7470A	PPB	PPB	17 Apr 2018 14:56:28	U	SMPL
J2370-08 - 1	CG	1061	0.3148	-7470A	PPB	PPB	17 Apr 2018 14:58:35	U	SMPL
J2386-09 - 1	PAINT-CHIPS-1	43	0.0371	-7470A	PPB	PPB	17 Apr 2018 15:00:43	U	SMPL
J2386-11 - 1	PAINT-CHIPS-2	39	0.0360	-7470A	PPB	PPB	17 Apr 2018 15:02:50	U	SMPL
PB108270BL - 1	PBW	-9	0.0229	-7470A	PPB	PPB	17 Apr 2018 15:04:57	U	SMPL
PB108270BS - 1	LCSW	13895	3.8156	-7470A	PPB	PPB	17 Apr 2018 15:07:05	U	SMPL
J2383-04 - 1	DRUM-PW-COMPOSITE	-36	0.0155	-7470A	PPB	PPB	17 Apr 2018 15:09:13	U	SMPL
J2383-04DUP - 1	DRUM-PW-COMPOSITEDUP	53	0.0398	-7470A	PPB	PPB	17 Apr 2018 15:11:21	U	SMPL
J2383-04MS - 1	DRUM-PW-COMPOSITEMS	14728	4.0429	-7470A	PPB	PPB	17 Apr 2018 15:13:28	U	SMPL
CCV56 - 1	CCV56	18153	4.9772	-7470A	PPB	PPB	17 Apr 2018 15:15:35	U	SMPL
CCB56 - 1	CCB56	-163	-0.0191	-7470A	PPB	PPB	17 Apr 2018 15:17:43	U	SMPL
J2383-04MSD - 1	DRUM-PW-COMPOSITEMS	14425	3.9602	-7470A	PPB	PPB	17 Apr 2018 15:19:51	U	SMPL
J2369-04LX5 - 1		-12	0.0221	-7470A	PPB	PPB	17 Apr 2018 15:24:05	U	SMPL
J2369-04A - 1		13748	3.7756	-7470A	PPB	PPB	17 Apr 2018 15:26:12	U	SMPL
J2383-04LX5 - 1		-120	-0.0074	-7470A	PPB	PPB	17 Apr 2018 15:28:20	U	SMPL
J2383-04A - 1		15290	4.1962	-7470A	PPB	PPB	17 Apr 2018 15:30:28	U	SMPL
CCV57 - 1	CCV57	17933	4.9171	-7470A	PPB	PPB	17 Apr 2018 15:34:45	U	SMPL
CCB57 - 1	CCB57	-175	-0.0224	-7470A	PPB	PPB	17 Apr 2018 15:36:53	U	SMPL