



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94683

Slope : 99.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	04/17/2018	15:40
2	CAL2	1	Water			20.7	7.00	04/17/2018	15:42
3	CAL3	1	Water			20.3	10.02	04/17/2018	15:44
4	ICV	1	Water			20.7	7.00	04/17/2018	15:46
5	CCV1	1	Water			20.5	2.00	04/17/2018	15:48
6	J2423-03	1	Solid	20.5	20	20.3	6.29	04/17/2018	15:50
7	J2423-03DUP	1	Solid	20.3	20	20.4	6.30	04/17/2018	15:52
8	J2423-06	1	Solid	20.5	20	20.7	6.85	04/17/2018	15:54
9	J2425-10	1	Solid	20.2	20	20.2	7.65	04/17/2018	15:56
10	J2427-10	1	Solid	20.3	20	20.5	7.16	04/17/2018	15:58
11	J2428-06	1	Solid	20.0	20	20.1	7.29	04/17/2018	16:00
12	CCV2	1	Water			20.5	12.00	04/17/2018	16:02

WORKLIST(Hardcopy Internal Chain)

LB 94683

WorkList Name : corrosivity J2428

WorkList ID : 110906

Date : 4/17/2018 3:25:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/17/2018	Solid	J2423-03	Corrosivity	Cool 4 deg C	PSEG01	G11	NYS-RBR200034	04/17/2018	9045C
04/17/2018	Solid	J2423-06	Corrosivity	Cool 4 deg C	PSEG01	G11	NYS-RBR200035	04/17/2018	9045C
04/17/2018	Solid	J2425-10	Corrosivity	Cool 4 deg C	PSEG01	G21	TP-3	04/17/2018	9045C
04/17/2018	Solid	J2427-10	Corrosivity	Cool 4 deg C	PSEG01	G21	TP-2	04/17/2018	9045C
04/17/2018	Solid	J2428-06	Corrosivity	Cool 4 deg C	PSEG01	G12	TP-1-TP-2	04/17/2018	9045C

Date/Time 04/17/18 3:35 PM

Received by: SP

Relinquished by: SP

Date/Time 04/17/18 4:10 PM

Received by: SP

Relinquished by: SP