


ISM02.4

LB94684
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40588



A= 0.0000e+000

B= 2.4740e-004

C= -3.7267e-002

Rho= 0.9999851

Accept=Accepted

Slope
 $y = mx + b$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	-0.011	-0.011	105	0.000	105					-
0.2	0.200	0.215	0.015	1019	0.0 %	1019					7
2.5	2.500	2.518	0.018	10327	0.0 %	10327					1
5.0	5.000	4.961	-0.039	20201	0.0 %	20201					-1
7.5	7.500	7.514	0.014	30522	0.0 %	30522					0
10.0	10.000	10.004	0.004	40588	0.0 %	40588					0

NB
4/17/18

I INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	105	-		0 ISM02.4	PPB	4/17/2018 15:50	Std
	0.2 S0.2	1019	-		0.2 ISM02.4	PPB	4/17/2018 15:53	Std
	2.5 S2.5	10327	-		2.5 ISM02.4	PPB	4/17/2018 15:55	Std
	5 S5	20201	-		5 ISM02.4	PPB	4/17/2018 15:57	Std
	7.5 S7.5	30522	-		7.5 ISM02.4	PPB	4/17/2018 16:00	Std
	10 S10.0	40588	-		10 ISM02.4	PPB	4/17/2018 16:02	Std
ICV	ICV	15965	3.9125 -		ISM02.4	PPB	4/17/2018 16:05	SMPL
ICB	ICB	-71	-0.0548 -		ISM02.4	PPB	4/17/2018 16:07	SMPL
CCV058	CCV058	20330	4.9924 -		ISM02.4	PPB	4/17/2018 16:09	SMPL
CCB058	CCB058	-68	-0.0541 -		ISM02.4	PPB	4/17/2018 16:12	SMPL
PB108295BL	PBW003	34	-0.0289 -		ISM02.4	PPB	4/17/2018 16:14	SMPL
J2329-01	MF1A30	353	0.0501 -		ISM02.4	PPB	4/17/2018 16:16	SMPL
J2329-02	MF1A30D	261	0.0273 -		ISM02.4	PPB	4/17/2018 16:18	SMPL
J2329-03	MF1A30S	3941	0.9377 -		ISM02.4	PPB	4/17/2018 16:21	SMPL
J2329-04	MF1A32	-43	-0.0479 -		ISM02.4	PPB	4/17/2018 16:23	SMPL
J2329-05	MF1A34	42	-0.0269 -		ISM02.4	PPB	4/17/2018 16:25	SMPL
J2329-06	MF1A48	183	0.008 -		ISM02.4	PPB	4/17/2018 16:27	SMPL
J2329-07	MF1A50	212	0.0152 -		ISM02.4	PPB	4/17/2018 16:30	SMPL
J2329-08	MF1A54	12	-0.0343 -		ISM02.4	PPB	4/17/2018 16:32	SMPL
J2329-09	MF1A56	25	-0.0311 -		ISM02.4	PPB	4/17/2018 16:34	SMPL
J2329-10	MF1A36	95	-0.0138 -		ISM02.4	PPB	4/17/2018 16:37	SMPL
J2329-11	MF1A38	104	-0.0115 -		ISM02.4	PPB	4/17/2018 16:39	SMPL
J2329-12	MF1A40	1275	0.2782 -		ISM02.4	PPB	4/17/2018 16:41	SMPL
J2329-13	MF1A42	12	-0.0343 -		ISM02.4	PPB	4/17/2018 16:43	SMPL
J2329-14	MF1A44	371	0.0545 -		ISM02.4	PPB	4/17/2018 16:46	SMPL
J2329-15	MF1A46	35	-0.0286 -		ISM02.4	PPB	4/17/2018 16:48	SMPL
J2329-16	MF1A52	154	0.0008 -		ISM02.4	PPB	4/17/2018 16:50	SMPL
J2329-17	MF1A57	49	-0.0251 -		ISM02.4	PPB	4/17/2018 16:52	SMPL
J2329-18	MF1A58	39	-0.0276 -		ISM02.4	PPB	4/17/2018 16:55	SMPL
J2329-19	MF1A59	32	-0.0294 -		ISM02.4	PPB	4/17/2018 16:57	SMPL
CCV059	CCV059	19864	4.8771 -		ISM02.4	PPB	4/17/2018 16:59	SMPL
CCB059	CCB059	-2	-0.0378 -		ISM02.4	PPB	4/17/2018 17:02	SMPL
PB108296BL	PBW003	18	-0.0328 -		ISM02.4	PPB	4/17/2018 17:04	SMPL
J2340-01	MF1A31	44	-0.0264 -		ISM02.4	PPB	4/17/2018 17:06	SMPL
J2340-02	MF1A31D	69	-0.0202 -		ISM02.4	PPB	4/17/2018 17:08	SMPL
J2340-03	MF1A31S	3668	0.8702 -		ISM02.4	PPB	4/17/2018 17:11	SMPL
J2340-04	MF1A33	-66	-0.0536 -		ISM02.4	PPB	4/17/2018 17:13	SMPL
J2340-05	MF1A35	16	-0.0333 -		ISM02.4	PPB	4/17/2018 17:15	SMPL
J2340-06	MF1A37	41	-0.0271 -		ISM02.4	PPB	4/17/2018 17:18	SMPL
J2340-07	MF1A39	34	-0.0289 -		ISM02.4	PPB	4/17/2018 17:20	SMPL
J2340-08	MF1A41	136	-0.0036 -		ISM02.4	PPB	4/17/2018 17:22	SMPL
J2340-09	MF1A55	11	-0.0345 -		ISM02.4	PPB	4/17/2018 17:24	SMPL
J2340-10	MF1A43	35	-0.0286 -		ISM02.4	PPB	4/17/2018 17:27	SMPL
J2340-11	MF1A45	12	-0.0343 -		ISM02.4	PPB	4/17/2018 17:29	SMPL
J2340-12	MF1A47	24	-0.0313 -		ISM02.4	PPB	4/17/2018 17:31	SMPL
J2340-13	MF1A60	15	-0.0336 -		ISM02.4	PPB	4/17/2018 17:34	SMPL

LB94684

I NSTRUMENT ID : CV1

CCV060	CCV060	19947	4.8977 -	ISM02.4	PPB	4/17/2018 17:36 SMPL
CCB060	CCB060	-21	-0.0425 -	ISM02.4	PPB	4/17/2018 17:38 SMPL