


ISM02.4

L1394691
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40077



A= 0.0000e+000

B= 2.4950e-004

C= -3.2693e-002

Rho= 0.9999196

Accept=Accepted

Slope
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	-0.009	-0.009	96	0.000	96					1
0.2	0.200	0.215	0.015	993	0.0 %	993					7
2.5	2.500	2.511	0.011	10194	0.0 %	10194					0
5.0	5.000	4.933	-0.067	19901	0.0 %	19901					1
7.5	7.500	7.584	0.084	30528	0.0 %	30528					1
10.0	10.000	9.966	-0.034	40077	0.0 %	40077					0

MB
4/17/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	96	-		0 ISM02.4	PPB	4/17/2018 17:52	Std
	0.2 S0.2	993	-		0.2 ISM02.4	PPB	4/17/2018 17:54	Std
	2.5 S2.5	10194	-		2.5 ISM02.4	PPB	4/17/2018 17:56	Std
	5 S5	19901	-		5 ISM02.4	PPB	4/17/2018 17:59	Std
	7.5 S7.5	30528	-		7.5 ISM02.4	PPB	4/17/2018 18:01	Std
	10 S10.0	40077	-		10 ISM02.4	PPB	4/17/2018 18:03	Std
ICV	ICV	16018	3.9638 -		ISM02.4	PPB	4/17/2018 18:06	SMPL
ICB	ICB	-41	-0.0429 -		ISM02.4	PPB	4/17/2018 18:08	SMPL
CCV061	CCV061	20345	5.0433 -		ISM02.4	PPB	4/17/2018 18:11	SMPL
CCB061	CCB061	-47	-0.0444 -		ISM02.4	PPB	4/17/2018 18:13	SMPL
PB108297BL	PBW010	72	-0.0147 -		ISM02.4	PPB	4/17/2018 18:15	SMPL
PB108297TB	LEB010	158	0.0067 -		ISM02.4	PPB	4/17/2018 18:17	SMPL
J2389-01	MESSM4	11	-0.0299 -		ISM02.4	PPB	4/17/2018 18:20	SMPL
J2389-02	MESSM4D	4	-0.0317 -		ISM02.4	PPB	4/17/2018 18:22	SMPL
J2389-03	MESSM4S	3856	0.9294 -		ISM02.4	PPB	4/17/2018 18:24	SMPL
CCV062	CCV062	19938	4.9418 -		ISM02.4	PPB	4/17/2018 18:26	SMPL
CCB062	CCB062	-46	-0.0442 -		ISM02.4	PPB	4/17/2018 18:29	SMPL