

L1394692


7471B

INSTRUMENT ID: CV4

Linear

μ Abs.:

43663



A= 0.0000e+000

B= 2.2875e-004

C= 1.3013e-002

Rho= 0.9998513

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	104	0.000	104				
0.2	0.200	0.231	0.031	951	0.0 %	951				
2.5	2.500	2.464	-0.036	10714	0.0 %	10714				
5.0	5.000	4.883	-0.117	21291	0.0 %	21291				
7.5	7.500	7.584	0.084	33098	0.0 %	33098				
10	10.000	10.001	0.001	43663	0.0 %	43663				

LB94692 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Apr 2018 18:40:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7471B	PPB	17 Apr 2018 18:41:40	S	Std
0.2 - 1	50.2	951	-	0.2000	7471B	PPB	17 Apr 2018 18:43:57	S	Std
2.5 - 1	52.5	10714	-	2.5000	7471B	PPB	17 Apr 2018 18:46:15	S	Std
5.0 - 1	55	21291	-	5.0000	7471B	PPB	17 Apr 2018 18:48:32	S	Std
7.5 - 1	57.5	33098	-	7.5000	7471B	PPB	17 Apr 2018 18:50:49	S	Std
10 - 1	510	43663	-	10.0000	7471B	PPB	17 Apr 2018 18:53:07	S	Std
ICV99 - 1	ICV99	17225	3.9533	-	7471B	PPB	17 Apr 2018 18:56:10	U	SMPL
ICB99 - 1	ICB99	43	0.0228	-	7471B	PPB	17 Apr 2018 18:58:26	U	SMPL
CCV63 - 1	CCV63	21344	4.8955	-	7471B	PPB	17 Apr 2018 19:00:41	U	SMPL
CCB63 - 1	CCB63	0	0.0130	-	7471B	PPB	17 Apr 2018 19:02:57	U	SMPL
CRA - 1	CRA	912	0.2216	-	7471B	PPB	17 Apr 2018 19:05:14	U	SMPL
HighStd - 1	HighStd	43780	10.0278	-	7471B	PPB	17 Apr 2018 19:07:30	U	SMPL
ChkStd - 1	ChkStd	30625	7.0186	-	7471B	PPB	17 Apr 2018 19:09:46	U	SMPL
PB108325BL - 1	PB108325BL	-123	-0.0151	-	7471B	PPB	17 Apr 2018 19:12:04	U	SMPL
PB108325BS - 1	PB108325BS	14368	3.2997	-	7471B	PPB	17 Apr 2018 19:14:22	U	SMPL
J2407-01 - 1	SPOILS-ICTF-A	21838	5.0085	-	7471B	PPB	17 Apr 2018 19:16:39	U	SMPL
J2407-03 - 1	SPOILS-ICTF-D	14612	3.3556	-	7471B	PPB	17 Apr 2018 19:18:56	U	SMPL
J2407-05 - 1	SPOILS-ICTF-E	16593	3.8087	-	7471B	PPB	17 Apr 2018 19:21:14	U	SMPL
J2408-01 - 1	NAA-WS-005	379	0.0997	-	7471B	PPB	17 Apr 2018 19:23:32	U	SMPL
J2408-01DUP - 1	NAA-WS-005DUP	401	0.1047	-	7471B	PPB	17 Apr 2018 19:25:48	U	SMPL
CCV64 - 1	CCV64	21395	4.9072	-	7471B	PPB	17 Apr 2018 19:28:04	U	SMPL
CCB64 - 1	CCB64	59	0.0265	-	7471B	PPB	17 Apr 2018 19:30:20	U	SMPL
J2408-01MS - 1	NAA-WS-005MS	16523	3.7927	-	7471B	PPB	17 Apr 2018 19:32:37	U	SMPL
J2408-01MSD - 1	NAA-WS-005MSD	16305	3.7428	-	7471B	PPB	17 Apr 2018 19:34:53	U	SMPL
J2408-02 - 1	NAA-WS-005	397	0.1038	-	7471B	PPB	17 Apr 2018 19:37:10	U	SMPL
J2410-01 - 1	T-2	1565	0.3710	-	7471B	PPB	17 Apr 2018 19:39:26	U	SMPL
J2410-05 - 1	T-3	168	0.0514	-	7471B	PPB	17 Apr 2018 19:41:43	U	SMPL
J2411-01 - 1	DRUM-1-3	1497	0.3555	-	7471B	PPB	17 Apr 2018 19:44:00	U	SMPL
J2408-01LX5 - 1		149	0.0471	-	7471B	PPB	17 Apr 2018 19:46:17	U	SMPL
J2408-01A - 1		17490	4.0139	-	7471B	PPB	17 Apr 2018 19:48:35	U	SMPL
CCV65 - 1	CCV65	21252	4.8745	-	7471B	PPB	17 Apr 2018 19:56:18	U	SMPL
CCB65 - 1	CCB65	-21	0.0082	-	7471B	PPB	17 Apr 2018 19:58:33	U	SMPL