

7470A


LB 94696

INSTRUMENT ID: LY1

Linear

 μ Abs.:

35186



A= 0.0000e+000

B= 2.8565e-004

C= -3.9360e-002

Rho= 0.9998589

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	63	0.000	63				
0.2	0.200	0.178	-0.022	762	0.0 %	762				
2.5	2.500	2.619	0.119	9305	0.0 %	9305				
5.0	5.000	4.911	-0.089	17330	0.0 %	17330				
7.5	7.500	7.502	0.002	26400	0.0 %	26400				
10.0	10.000	10.012	0.012	35186	0.0 %	35186				

LB94696 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Apr 2018 10:25:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.0000	7470A	PPB	18 Apr 2018 10:31:45	S	Std
0.2 - 1	50.2	762	-	0.2000	7470A	PPB	18 Apr 2018 10:33:53	S	Std
2.5 - 1	52.5	9305	-	2.5000	7470A	PPB	18 Apr 2018 10:36:02	S	Std
5.0 - 1	55	17330	-	5.0000	7470A	PPB	18 Apr 2018 10:38:10	S	Std
7.5 - 1	57.5	26400	-	7.5000	7470A	PPB	18 Apr 2018 10:40:18	S	Std
10.0 - 1	510	35186	-	10.0000	7470A	PPB	18 Apr 2018 10:44:56	S	Std
ICV01 - 1	ICV01	13488	3.8135	-	7470A	PPB	18 Apr 2018 10:47:30	U	SMPL
ICB01 - 1	ICB01	-98	-0.0674	-	7470A	PPB	18 Apr 2018 10:49:36	U	SMPL
CCV66 - 1	CCV66	17545	4.9724	-	7470A	PPB	18 Apr 2018 10:51:43	U	SMPL
CCB66 - 1	CCB66	-103	-0.0688	-	7470A	PPB	18 Apr 2018 10:53:50	U	SMPL
CRA - 1	CRA	748	0.1743	-	7470A	PPB	18 Apr 2018 10:55:57	U	SMPL
HighStd - 1	HighStd	35243	10.0279	-	7470A	PPB	18 Apr 2018 10:58:04	U	SMPL
ChkStd - 1	ChkStd	23660	6.7192	-	7470A	PPB	18 Apr 2018 11:00:11	U	SMPL
PB108330BL - 1	PBW	28	-0.0314	-	7470A	PPB	18 Apr 2018 11:05:14	U	SMPL
PB108330BS - 1	LCSW	13484	3.8124	-	7470A	PPB	18 Apr 2018 11:07:21	U	SMPL
PB108330TB - 1	PB108299TB	-103	-0.0688	-	7470A	PPB	18 Apr 2018 11:09:29	U	SMPL
J2410-04 - 1	T-2	-21	-0.0454	-	7470A	PPB	18 Apr 2018 11:11:37	U	SMPL
J2410-08 - 1	T-3	0	-0.0394	-	7470A	PPB	18 Apr 2018 11:13:46	U	SMPL
J2411-03 - 1	DRUM-1-3	40	-0.0279	-	7470A	PPB	18 Apr 2018 11:15:54	U	SMPL
J2411-03DUP - 1	DRUM-1-3DUP	-17	-0.0442	-	7470A	PPB	18 Apr 2018 11:18:01	U	SMPL
CCV67 - 1	CCV67	17535	4.9695	-	7470A	PPB	18 Apr 2018 11:20:07	U	SMPL
CCB67 - 1	CCB67	-99	-0.0676	-	7470A	PPB	18 Apr 2018 11:22:14	U	SMPL
J2411-03MS - 1	DRUM-1-3MS	13338	3.7707	-	7470A	PPB	18 Apr 2018 11:24:21	U	SMPL
J2411-03MSD - 1	DRUM-1-3MSD	13131	3.7115	-	7470A	PPB	18 Apr 2018 11:26:28	U	SMPL
PB108331BL - 1	PBW	-94	-0.0662	-	7470A	PPB	18 Apr 2018 11:28:35	U	SMPL
PB108331BS - 1	LCSW	11441	3.2288	-	7470A	PPB	18 Apr 2018 11:30:43	U	SMPL
PB108331TB - 1	PB108282TB	-85	-0.0636	-	7470A	PPB	18 Apr 2018 11:32:51	U	SMPL
J2387-11 - 1	20180244-AMEND-COMP-A	478	0.0972	-	7470A	PPB	18 Apr 2018 11:34:59	U	SMPL
J2387-11DUP - 1	20180244-AMEND-COMP-ADUP	479	0.0975	-	7470A	PPB	18 Apr 2018 11:37:07	U	SMPL
J2387-11MS - 1	20180244-AMEND-COMP-AMS	14483	4.0977	-	7470A	PPB	18 Apr 2018 11:39:15	U	SMPL
J2387-11MSD - 1	20180244-AMEND-COMP-AMSD	14210	4.0198	-	7470A	PPB	18 Apr 2018 11:41:23	U	SMPL
J2387-12 - 1	20180245-AMEND-COMP-B	377	0.0683	-	7470A	PPB	18 Apr 2018 11:43:30	U	SMPL
CCV68 - 1	CCV68	17472	4.9516	-	7470A	PPB	18 Apr 2018 11:45:38	U	SMPL
CCB68 - 1	CCB68	-54	-0.0548	-	7470A	PPB	18 Apr 2018 11:47:44	U	SMPL
J2387-13 - 1	20180246-AMEND-COMP-C	582	0.1269	-	7470A	PPB	18 Apr 2018 11:49:52	U	SMPL
J2411-03LX5 - 1		123	-0.0042	-	7470A	PPB	18 Apr 2018 11:51:59	U	SMPL
J2411-03A - 1		15375	4.3525	-	7470A	PPB	18 Apr 2018 11:54:06	U	SMPL
J2387-11LX5 - 1		228	0.0258	-	7470A	PPB	18 Apr 2018 11:56:13	U	SMPL
J2387-11A - 1		14589	4.1280	-	7470A	PPB	18 Apr 2018 11:58:20	U	SMPL
CCV69 - 1	CCV69	17833	5.0547	-	7470A	PPB	18 Apr 2018 12:00:27	U	SMPL
CCB69 - 1	CCB69	-10	-0.0422	-	7470A	PPB	18 Apr 2018 12:02:36	U	SMPL