

7471B


LB94699

INSTRUMENT ID: CY1

Linear

 μ Abs.:

44273



A= 0.0000e+000

B= 2.2294e-004

C= -2.2457e-002

Rho= 0.9989783

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	103	0.000	103				
0.2	0.200	0.204	0.004	1017	0.0 %	1017				
2.5	2.500	2.559	0.059	11578	0.0 %	11578				
5.0	5.000	4.783	-0.217	21555	0.0 %	21555				
7.5	7.500	7.805	0.305	35111	0.0 %	35111				
10	10.000	9.848	-0.152	44273	0.0 %	44273				

LB94699 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Apr 2018 12:28:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	103	-	0.0000	7471B	PPB	18 Apr 2018 12:30:27	S	Std
0.2 - 1	50.2	1017	-	0.2000	7471B	PPB	18 Apr 2018 12:32:44	S	Std
2.5 - 1	52.5	11578	-	2.5000	7471B	PPB	18 Apr 2018 12:35:01	S	Std
5.0 - 1	55	21555	-	5.0000	7471B	PPB	18 Apr 2018 12:37:19	S	Std
7.5 - 1	57.5	35111	-	7.5000	7471B	PPB	18 Apr 2018 12:39:36	S	Std
10 - 1	510	44273	-	10.0000	7471B	PPB	18 Apr 2018 12:41:53	S	Std
ICV02 - 1	ICV02	17006	3.7689	-7471B	PPB	PPB	18 Apr 2018 12:44:59	U	SMPL
ICB02 - 1	ICB02	-53	-0.0343	-7471B	PPB	PPB	18 Apr 2018 12:47:15	U	SMPL
CCV70 - 1	CCV70	21599	4.7929	-7471B	PPB	PPB	18 Apr 2018 12:49:31	U	SMPL
CCB70 - 1	CCB70	-83	-0.0410	-7471B	PPB	PPB	18 Apr 2018 12:51:47	U	SMPL
CRA - 1	CRA	984	0.1969	-7471B	PPB	PPB	18 Apr 2018 12:54:04	U	SMPL
HighStd - 1	HighStd	44209	9.8337	-7471B	PPB	PPB	18 Apr 2018 12:56:19	U	SMPL
ChkStd - 1	ChkStd	32943	7.3220	-7471B	PPB	PPB	18 Apr 2018 12:58:35	U	SMPL
PB108337BL - 1	PBS	-159	-0.0579	-7471B	PPB	PPB	18 Apr 2018 13:00:54	U	SMPL
PB108337BS - 1	LCSS	15419	3.4151	-7471B	PPB	PPB	18 Apr 2018 13:03:12	U	SMPL
J2413-01 - 1	COMP-B1-0-6	2112	0.4484	-7471B	PPB	PPB	18 Apr 2018 13:05:29	U	SMPL
J2413-02 - 1	COMP-B1-18-20	224	0.0275	-7471B	PPB	PPB	18 Apr 2018 13:07:46	U	SMPL
J2413-03 - 1	COMP-B2-0-6	24414	5.4205	-7471B	PPB	PPB	18 Apr 2018 13:10:03	U	SMPL
J2413-04 - 1	COMP-B2-18-20	169	0.0152	-7471B	PPB	PPB	18 Apr 2018 13:12:21	U	SMPL
J2423-01 - 1	NYS-RBR200034	2010	0.4257	-7471B	PPB	PPB	18 Apr 2018 13:14:38	U	SMPL
CCV71 - 1	CCV71	21167	4.6966	-7471B	PPB	PPB	18 Apr 2018 13:16:54	U	SMPL
CCB71 - 1	CCB71	-56	-0.0349	-7471B	PPB	PPB	18 Apr 2018 13:19:10	U	SMPL
J2423-04 - 1	NYS-RBR200035	1626	0.3401	-7471B	PPB	PPB	18 Apr 2018 13:21:26	U	SMPL
J2424-01 - 1	FK-GF-02-016-A	2390	0.5104	-7471B	PPB	PPB	18 Apr 2018 13:23:43	U	SMPL
J2424-03 - 1	FK-GF-02-016-B	2034	0.4310	-7471B	PPB	PPB	18 Apr 2018 13:25:59	U	SMPL
J2425-01 - 1	T-3	763	0.1476	-7471B	PPB	PPB	18 Apr 2018 13:28:15	U	SMPL
J2427-01 - 1	T-2	382	0.0627	-7471B	PPB	PPB	18 Apr 2018 13:30:31	U	SMPL
J2427-01DUP - 1	T-2DUP	494	0.0877	-7471B	PPB	PPB	18 Apr 2018 13:32:49	U	SMPL
J2427-01MS - 1	T-2MS	16746	3.7110	-7471B	PPB	PPB	18 Apr 2018 13:35:06	U	SMPL
J2427-01MSD - 1	T-2MSD	17132	3.7970	-7471B	PPB	PPB	18 Apr 2018 13:37:24	U	SMPL
J2428-01 - 1	TP-1-TP-2	3170	0.6843	-7471B	PPB	PPB	18 Apr 2018 13:39:42	U	SMPL
CCV72 - 1	CCV72	21332	4.7334	-7471B	PPB	PPB	18 Apr 2018 13:41:58	U	SMPL
CCB72 - 1	CCB72	-114	-0.0479	-7471B	PPB	PPB	18 Apr 2018 13:44:14	U	SMPL
J2427-01LX5 - 1		112	0.0025	-7471B	PPB	PPB	18 Apr 2018 13:46:32	U	SMPL
J2427-01A - 1		18775	4.1633	-7471B	PPB	PPB	18 Apr 2018 13:51:32	U	SMPL
CCV73 - 1	CCV73	21667	4.8081	-7471B	PPB	PPB	18 Apr 2018 13:53:48	U	SMPL
CCB73 - 1	CCB73	-95	-0.0436	-7471B	PPB	PPB	18 Apr 2018 13:56:05	U	SMPL