



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94704

Slope : 99.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	04/18/2018	11:40
2	CAL2	1	Water			20.3	7.00	04/18/2018	11:42
3	CAL3	1	Water			20.5	10.03	04/18/2018	11:44
4	ICV	1	Water			20.2	7.00	04/18/2018	11:46
5	CCV1	1	Water			20.5	2.00	04/18/2018	11:48
6	J2433-01	1	Water			18.4	4.61	04/18/2018	11:50
7	J2433-01DUP	1	Water			18.6	4.62	04/18/2018	11:52
8	CCV2	1	Water			20.5	12.00	04/18/2018	11:54

WORKLIST(Hardcopy Internal Chain)

LB 94704

Worklist Name : pHJ2433

Worklist ID : 110935

Date : 4/18/2018 11:28:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/11/2018	Water	J2433-01	pH	Cool 4 deg C	STER10	G11	CELL-A-POOL-WATER(DE-1C	04/11/2018	SM 4500-PH B

Date/Time 04/18/18 11:35 AM

Received by: SP

Relinquished by: CP

Date/Time 04/18/18 12:05

Received by: CP

Relinquished by: SP