



## Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94712

Slope : 97.8

pH Meter ID : wc ph meter-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2271         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2220         |
| buffer solution pH 7 yellow    | W2219         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 21.3             | 4.01        | 04/18/2018 | 16:25     |
| 2   | CAL2        | 1  | Water  |             |             | 21.3             | 7.00        | 04/18/2018 | 16:26     |
| 3   | CAL3        | 1  | Water  |             |             | 21.3             | 10.02       | 04/18/2018 | 16:27     |
| 4   | ICV         | 1  | Water  |             |             | 21.4             | 7.01        | 04/18/2018 | 16:30     |
| 5   | CCV1        | 1  | Water  |             |             | 21.3             | 2.01        | 04/18/2018 | 16:31     |
| 6   | J2454-01    | 1  | Water  |             |             | 20.1             | 7.80        | 04/18/2018 | 16:32     |
| 7   | J2454-01DUP | 1  | Water  |             |             | 20.2             | 7.81        | 04/18/2018 | 16:34     |
| 8   | CCV2        | 1  | Water  |             |             | 21.3             | 12.01       | 04/18/2018 | 16:35     |

# WORKLIST(Hardcopy Internal Chain)

1344+12

WorkList Name : PH J2434

WorkList ID : 110948

Date : 4/18/2018 4:17:41 PM

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|-------------|------|--------------|----------|------------------|-----------------|--------------|--------------|
| 04/17/2018 | Water  | J2454-01-17 | pH   | Cool 4 deg C | HOME01   | G21              | PARAMUS         | 04/17/2018   | SM 4500-PH B |

Date/Time 04-18-18 4:20 PM  
 Received by: JP  
 Relinquished by: CJ

Date/Time 04-18-18 5:15 PM  
 Received by: CJ  
 Relinquished by: JP