



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94713

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	04/18/2018	16:25
2	CAL2	1	Water			21.3	7.00	04/18/2018	16:26
3	CAL3	1	Water			21.3	10.02	04/18/2018	16:27
4	ICV	1	Water			21.4	7.01	04/18/2018	16:30
5	CCV1	1	Water			21.2	2.02	04/18/2018	16:39
6	J2438-01	1	Solid	20.02	20	20.4	9.52	04/18/2018	16:40
7	J2438-01DUP	1	Solid	20.03	20	20.2	9.53	04/18/2018	16:41
8	CCV2	1	Water			21.3	12.01	04/18/2018	16:42

1394713

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2438

WorkList ID : 110949

Date : 4/18/2018 4:18:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/13/2018	Solid	J2438-01 -1A	pH	Cool 4 deg C	USEP01	A11	MH0S17	04/13/2018	9045C

Date/Time 04-18-18 4:20 PM
Received by: SP
Relinquished by: CP

Date/Time 04-18-18 5:25 PM
Received by: CP
Relinquished by: JP