

LB94723


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35375



A= 0.0000e+000

B= 2.8231e-004

C= 2.9768e-002

Rho= 0.9997813

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	-20	0.000	-20				
0.2	0.200	0.227	0.027	698	0.0 %	698				
2.5	2.500	2.523	0.023	8833	0.0 %	8833				
5.0	5.000	4.833	-0.167	17015	0.0 %	17015				
7.5	7.500	7.576	0.076	26731	0.0 %	26731				
10.0	10.000	10.016	0.016	35375	0.0 %	35375				

LB94723 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Apr 2018 08:56:49

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-20	-	0.0000	7470A	PPB	19 Apr 2018 09:00:08	S	Std
0.2 - 1	50.2	698	-	0.2000	7470A	PPB	19 Apr 2018 09:02:16	S	Std
2.5 - 1	52.5	8833	-	2.5000	7470A	PPB	19 Apr 2018 09:04:25	S	Std
5.0 - 1	55	17015	-	5.0000	7470A	PPB	19 Apr 2018 09:06:33	S	Std
7.5 - 1	57.5	26731	-	7.5000	7470A	PPB	19 Apr 2018 09:08:41	S	Std
10.0 - 1	510	35375	-	10.0000	7470A	PPB	19 Apr 2018 09:10:49	S	Std
ICV03 - 1	ICV03	13373	3.8051	-7470A	PPB	19 Apr 2018 09:13:25	U	SMPL	
ICB03 - 1	ICB03	-156	-0.0143	-7470A	PPB	19 Apr 2018 09:15:31	U	SMPL	
CCV74 - 1	CCV74	17465	4.9603	-7470A	PPB	19 Apr 2018 09:17:38	U	SMPL	
CCB74 - 1	CCB74	-172	-0.0188	-7470A	PPB	19 Apr 2018 09:19:45	U	SMPL	
CRA - 1	CRA	724	0.2342	-7470A	PPB	19 Apr 2018 09:21:52	U	SMPL	
HighStd - 1	HighStd	35686	10.1042	-7470A	PPB	19 Apr 2018 09:23:59	U	SMPL	
ChkStd - 1	ChkStd	24694	7.0010	-7470A	PPB	19 Apr 2018 09:26:06	U	SMPL	
PB108373BL - 1	PBW	-240	-0.0380	-7470A	PPB	19 Apr 2018 09:28:15	U	SMPL	
PB108373BS - 1	LCSW	12712	3.6184	-7470A	PPB	19 Apr 2018 09:30:25	U	SMPL	
PB108373TB - 1	PB108334TB	-127	-0.0061	-7470A	PPB	19 Apr 2018 09:32:32	U	SMPL	
J2423-03 - 1	NYS-RBR200034	-32	0.0207	-7470A	PPB	19 Apr 2018 09:34:40	U	SMPL	
J2423-03DUP - 1	NYS-RBR200034DUP	-13	0.0261	-7470A	PPB	19 Apr 2018 09:36:48	U	SMPL	
J2423-03MS - 1	NYS-RBR200034MS	13520	3.8466	-7470A	PPB	19 Apr 2018 09:38:57	U	SMPL	
J2423-03MSD - 1	NYS-RBR200034MSD	13213	3.7599	-7470A	PPB	19 Apr 2018 09:41:04	U	SMPL	
CCV75 - 1	CCV75	17178	4.8792	-7470A	PPB	19 Apr 2018 09:43:11	U	SMPL	
CCB75 - 1	CCB75	-152	-0.0131	-7470A	PPB	19 Apr 2018 09:45:17	U	SMPL	
J2423-06 - 1	NYS-RBR200035	-37	0.0193	-7470A	PPB	19 Apr 2018 09:47:25	U	SMPL	
J2425-10 - 1	T-3	-15	0.0255	-7470A	PPB	19 Apr 2018 09:49:32	U	SMPL	
J2427-10 - 1	T-2	-14	0.0258	-7470A	PPB	19 Apr 2018 09:51:38	U	SMPL	
J2428-06 - 1	TP-1-TP-2	-10	0.0269	-7470A	PPB	19 Apr 2018 09:53:45	U	SMPL	
J2423-03LX5 - 1		-7	0.0278	-7470A	PPB	19 Apr 2018 09:55:52	U	SMPL	
J2423-03A - 1		13566	3.8595	-7470A	PPB	19 Apr 2018 09:58:00	U	SMPL	
CCV76 - 1	CCV76	17523	4.9766	-7470A	PPB	19 Apr 2018 10:00:08	U	SMPL	
CCB76 - 1	CCB76	-158	-0.0148	-7470A	PPB	19 Apr 2018 10:02:17	U	SMPL	