



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94725

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	04/19/2018	08:35
2	CAL2	1	Water			21.5	7.01	04/19/2018	08:31
3	CAL3	1	Water			21.6	10.03	04/19/2018	08:32
4	ICV	1	Water			21.5	6.99	04/19/2018	08:34
5	CCV1	1	Water			21.4	2.01	04/19/2018	08:35
6	J2157-12	1	Solid	20.03	20	20.5	8.70	04/19/2018	08:36
7	J2157-12DUP	1	Solid	20.05	20	20.5	8.71	04/19/2018	08:37
8	J2165-06	1	Solid	20.03	20	20.5	8.13	04/19/2018	08:39
9	J2439-03	1	Solid	20.04	20	19.8	8.12	04/19/2018	08:40
10	CCV2	1	Water			21.3	12.01	04/19/2018	08:41

2894725

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J12457 WorkList ID : 110977 Date : 4/19/2018 8:26:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/18/2018	Solid	J2157-12	Corrosivity	Cool 4 deg C	PSEG05		CL-02-041818	04/18/2018	9045C
04/18/2018	Solid	J2165-06	Corrosivity	Cool 4 deg C	PSEG05		SU-01-041818	04/18/2018	9045C
04/17/2018	Solid	J2439-03	Corrosivity	Cool 4 deg C	DIST01	G21	CDMP-WC1	04/17/2018	9045C

Date/Time 04-19-18 8:30 AM
Received by: JP
Relinquished by: CD

Date/Time 04-19-18 9:25 AM
Received by: CD
Relinquished by: SD