

ISM02.4


LB 94729

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45645



A= 0.0000e+000

B= 2.1880e-004

C= 6.5201e-003

Rho= 0.9999707

Accept= Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.016	0.016	42	0.000	42					-
0.2	0.200	0.220	0.020	977	0.0 %	977					10
2.5	2.500	2.460	-0.040	11215	0.0 %	11215					-2
5.0	5.000	4.969	-0.031	22682	0.0 %	22682					-1
7.5	7.500	7.540	0.040	34430	0.0 %	34430					1
10.0	10.000	9.995	-0.005	45649	0.0 %	45649					0

LB
4/19/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	42	-		0 ISM02.4	PPB	4/19/2018 10:35	Std
	0.2 S0.2	977	-		0.2 ISM02.4	PPB	4/19/2018 10:37	Std
	2.5 S2.5	11215	-		2.5 ISM02.4	PPB	4/19/2018 10:39	Std
	5 S5	22682	-		5 ISM02.4	PPB	4/19/2018 10:41	Std
	7.5 S7.5	34430	-		7.5 ISM02.4	PPB	4/19/2018 10:44	Std
	10 S10.0	45649	-		10 ISM02.4	PPB	4/19/2018 10:46	Std
ICV	ICV	17762	3.8928 -		ISM02.4	PPB	4/19/2018 10:49	SMPL
ICB	ICB	-101	-0.0156 -		ISM02.4	PPB	4/19/2018 10:51	SMPL
CCV077	CCV077	22481	4.9254 -		ISM02.4	PPB	4/19/2018 10:53	SMPL
CCB077	CCB077	-115	-0.0186 -		ISM02.4	PPB	4/19/2018 10:56	SMPL
PB108370BL	PBS004	23	0.0116 -		ISM02.4	PPB	4/19/2018 10:58	SMPL
J2438-01	MH0S17	278	0.0673 -		ISM02.4	PPB	4/19/2018 11:05	SMPL
J2438-02	MH0S17D	246	0.0603 -		ISM02.4	PPB	4/19/2018 11:07	SMPL
J2438-03	MH0S17S	11237	2.4652 -		ISM02.4	PPB	4/19/2018 11:10	SMPL
CCV078	CCV078	22564	4.9435 -		ISM02.4	PPB	4/19/2018 11:12	SMPL
CCB078	CCB078	-109	-0.0173 -		ISM02.4	PPB	4/19/2018 11:14	SMPL