



Analytical Summary Report

Analysis Method: 9045C
Parameter: pH
Run Number: LB94730
BalanceID: WC SC-4

Analyst By : Jignesh
Supervisor Review By : apatel
Slope : 97.5
pH Meter ID : wc ph meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.01	04/19/2018	09:00
2	CAL2	1	Water			21.2	7.00	04/19/2018	09:01
3	CAL3	1	Water			21.2	10.03	04/19/2018	09:02
4	ICV	1	Water			21.3	6.99	04/19/2018	09:05
5	CCV1	1	Water			21.4	2.01	04/19/2018	09:06
6	J2457-01	1	Solid	20.02	20	21.3	8.09	04/19/2018	09:09
7	J2457-01DUP	1	Solid	20.03	20	21.4	8.10	04/19/2018	09:11
8	J2457-04	1	Solid	20.02	20	21.1	8.16	04/19/2018	09:15
9	J2457-05	1	Solid	20.03	20	21.0	8.10	04/19/2018	09:16
10	J2457-06	1	Solid	20.04	20	20.9	8.22	04/19/2018	09:18
11	CCV2	1	Water			21.3	12.01	04/19/2018	09:20

21394730

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2457

WorkList ID : 110976

Date : 4/19/2018 8:25:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/18/2018	Solid	J2457-01 -1A	pH	Cool 4 deg C	ROYF02		EME-FW-TS001-01	04/18/2018	9045C
04/18/2018	Solid	J2457-04 -1A	pH	Cool 4 deg C	ROYF02		EME-FW-TS001-02	04/18/2018	9045C
04/18/2018	Solid	J2457-05 -1A	pH	Cool 4 deg C	ROYF02		EME-FW-TS002-01	04/18/2018	9045C
04/18/2018	Solid	J2457-06 -1A	pH	Cool 4 deg C	ROYF02		EME-FW-TS003-01	04/18/2018	9045C

Date/Time 04-14-18 8:30 AM
 Received by: JO
 Relinquished by: CB

Date/Time 04-19-18 9:30 AM
 Received by: CB
 Relinquished by: JO