



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94732

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP63841
Turbidity std, 40NTU	WP63835
turbidity std 5 NTU	WP63838
turbidity std 20 NTU	WP63836
Turbidity std 1 NTU	WP63839
turbidity std 0 NTU	WP63840
turbidity std 10 NTU	WP63837
10 NTU Standard 500 ml	W2242

Intercept: -0.0253

Slope: 0.9117

Regression: 0.9996

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.195	04/19/2018	09:22
2	CAL2	1	1	1.080	04/19/2018	09:26
3	CAL3	5	1	4.680	04/19/2018	09:31
4	CAL4	10	1	8.780	04/19/2018	09:35
5	CAL5	20	1	17.600	04/19/2018	09:39
6	CAL6	40	1	36.800	04/19/2018	09:43



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94732

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.610	10.568	04/19/2018	09:47
2	ICB		1	0.190	0.236	04/19/2018	09:51
3	CCV1	10	1	9.100	10.009	04/19/2018	09:54
4	CCB1		1	0.187	0.233	04/19/2018	09:57
5	LB94732BL		1	0.188	0.234	04/19/2018	10:00
6	J2454-01		1	0.242	0.293	04/19/2018	10:05
7	J2454-01DUP		1	0.240	0.291	04/19/2018	10:08
8	CCV2	10	1	8.990	9.888	04/19/2018	10:11
9	CCB2		1	0.187	0.233	04/19/2018	10:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBIDITY041918

WorkList ID : 110990

Date : 4/19/2018 9:01:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/19/2018	Water	J2454-01	Turbidity	Cool 4 deg C	HOME01	G21	PARAMUS	04/17/2018	SM2130 B

Date/Time 4-19-18 9:08
 Received by: AK
 Relinquished by: AK

Date/Time 4-19-18 10:30
 Received by: AK
 Relinquished by: AK