

245.1


LB94746

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34198



A= 0.0000e+000

B= 2.9315e-004

C= -1.1850e-002

Rho= 0.9999879

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	8	0.000	8				
0.05	0.050	0.043	-0.007	187	0.0 %	187				
0.2	0.200	0.204	0.004	738	0.0 %	738				
2.5	2.500	2.515	0.015	8620	0.0 %	8620				
5	5.000	5.020	0.020	17166	0.0 %	17166				
7.5	7.500	7.463	-0.037	25499	0.0 %	25499				
10	10.000	10.013	0.013	34198	0.0 %	34198				

LB94746 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Apr 2018 16:25:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	245.1	PPB	19 Apr 2018 16:27:20	S	Std
0.05 - 1	50.05	187	-	0.0500	245.1	PPB	19 Apr 2018 16:29:28	S	Std
0.2 - 1	50.2	738	-	0.2000	245.1	PPB	19 Apr 2018 16:31:36	S	Std
2.5 - 1	52.5	8620	-	2.5000	245.1	PPB	19 Apr 2018 16:35:31	S	Std
5 - 1	55	17166	-	5.0000	245.1	PPB	19 Apr 2018 16:37:39	S	Std
7.5 - 1	57.5	25499	-	7.5000	245.1	PPB	19 Apr 2018 16:39:46	S	Std
10 - 1	59	34198	-	10.0000	245.1	PPB	19 Apr 2018 16:44:33	S	Std
ICV05 - 1	ICV05	13331	3.8961	-245.1		PPB	19 Apr 2018 16:52:00	U	SMPL
ICB05 - 1	ICB05	-119	-0.0467	-245.1		PPB	19 Apr 2018 16:54:06	U	SMPL
CCV83 - 1	CCV83	17378	5.0825	-245.1		PPB	19 Apr 2018 16:56:13	U	SMPL
CCB83 - 1	CCB83	-130	-0.0500	-245.1		PPB	19 Apr 2018 16:58:20	U	SMPL
CRA - 1	CRA	230	0.0556	-245.1		PPB	19 Apr 2018 17:00:27	U	SMPL
HighStd - 1	HighStd	34545	10.1151	-245.1		PPB	19 Apr 2018 17:02:34	U	SMPL
ChkStd - 1	ChkStd	23597	6.9056	-245.1		PPB	19 Apr 2018 17:04:41	U	SMPL
PB108405BL - 1	PBW	-33	-0.0215	-245.1		PPB	19 Apr 2018 17:09:14	U	SMPL
PB108405BS - 1	LCSW	12978	3.7927	-245.1		PPB	19 Apr 2018 17:11:22	U	SMPL
J2133-10 - 1	LOD-MDL-WATER-04-QT2-2018	277	0.0694	-245.1		PPB	19 Apr 2018 17:16:10	U	SMPL
J2133-11 - 1	LOQ-WATER-05-QT2-2018	1399	0.3983	-245.1		PPB	19 Apr 2018 17:18:17	U	SMPL
J2133-12 - 1	MDL-WATER-06-QT2-2018	194	0.0450	-245.1		PPB	19 Apr 2018 17:20:25	U	SMPL
J2454-01 - 1	PARAMUS	-53	-0.0274	-245.1		PPB	19 Apr 2018 17:22:34	U	SMPL
CCV84 - 1	CCV84	17108	5.0034	-245.1		PPB	19 Apr 2018 17:24:41	U	SMPL
CCB84 - 1	CCB84	-23	-0.0186	-245.1		PPB	19 Apr 2018 17:29:25	U	SMPL