



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94760

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	3.99	04/20/2018	07:50
2	CAL2	1	Water			21.1	7.01	04/20/2018	07:51
3	CAL3	1	Water			21.2	10.02	04/20/2018	07:53
4	ICV	1	Water			21.3	7.00	04/20/2018	08:00
5	CCV1	1	Water			21.2	2.01	04/20/2018	08:01
6	J2477-01	1	Solid	20.03	20	20.5	4.60	04/20/2018	08:03
7	J2477-01DUP	1	Solid	20.04	20	20.6	4.61	04/20/2018	08:05
8	CCV2	1	Water			21.2	12.02	04/20/2018	08:06

2BQU760

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J2477

WorkList ID : 111013

Date : 4/20/2018 7:44:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/19/2018	Solid	J2477-01-A	pH	Cool 4 deg C	TULL02	G52	FK-PS-01-022	04/19/2018	9045C

Date/Time 04/20/18 7:50 AM
Received by: JL
Relinquished by: ab

Date/Time 04/20/18 8:32 AM
Received by: CR
Relinquished by: JP