



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94763

Ambient Barometric Pressure (mmHg): 765

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	78.9	04/20/2018	09:10
2	J2472-04		1	>100.0	>212.0	>212.0	04/20/2018	09:35
3	J2472-04DUP		1	>100.0	>212.0	>212.0	04/20/2018	10:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB 94763

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-042018

WorkList ID : 111028

Date : 4/20/2018 8:24:14 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/02/2018	Water	J2472-04 C	Flash Point	Cool 4 deg C	DVIR01	G22	TANK-2	04/18/2018	1010A

Date/Time 04/20/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 04/20/18 10:05 AM
Received by: CP
Relinquished by: KS