

=====

Test results

Aquakem 7.2AQ1

1394764
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/20/2018 11:04

Test: Residue
Total CN ug/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	99.089	0.0	0.098	
ICB1	0.025	0.0	0.005	
CCV1	245.819	0.0	0.236	
CCB1	-0.572	0.0	0.004	
PB108423BL	-0.716	0.0	0.004	
J2472-04	-0.556	0.0	0.004	
J2472-04DUP	-1.255	0.0	0.004	
CCV2	252.337	0.0	0.243	
CCB2	-1.031	0.0	0.004	

N	9
Mean	65.904
SD	108.8765
CV%	165.20

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Apr 20 10:01:36 2018

Fri Apr 20 11:03:56 2018

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.9087	µg/l	4/20/2018 10:20:16	
5.0PPBCN	A	Reactive CN	P	4.3868	µg/l	4/20/2018 10:20:17	
10PPBCN	A	Reactive CN	P	9.3887	µg/l	4/20/2018 10:20:18	
50PPBCN	A	Reactive CN	P	50.2658	µg/l	4/20/2018 10:20:19	
100PPBCN	A	Reactive CN	P	99.4626	µg/l	4/20/2018 10:20:20	
250PPBCN	A	Reactive CN	P	254.6109	µg/l	4/20/2018 10:20:21	
500PPBCN	A	Reactive CN	P	497.7938	µg/l	4/20/2018 10:20:22	
ICV1	S	Reactive CN	P	99.0885	µg/l	4/20/2018 11:03:48	
ICB1	S	Reactive CN	P	0.0251	µg/l	4/20/2018 11:03:49	
CCV1	S	Reactive CN	P	245.8189	µg/l	4/20/2018 11:03:50	
CCB1	S	Reactive CN	P	-0.5715	µg/l	4/20/2018 11:03:51	
PB108423BL	S	Reactive CN	P	-0.7165	µg/l	4/20/2018 11:03:52	
J2472-04	S	Reactive CN	P	-0.5564	µg/l	4/20/2018 11:03:53	
J2472-04DUP	S	Reactive CN	P	-1.2554	µg/l	4/20/2018 11:03:54	
CCV2	S	Reactive CN	P	252.3365	µg/l	4/20/2018 11:03:55	
CCB2	S	Reactive CN	P	-1.0312	µg/l	4/20/2018 11:03:56	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/20/2018 10:28


Test Total CN

Accepted 4/20/2018 10:28

Factor 1061
Bias 0.005

Coeff. of det. 0.999863

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9087	0.0000	
2	5.0PPBCN	0.009	4.3868	5.0000	
3	10PPBCN	0.014	9.3887	10.0000	
4	50PPBCN	0.052	50.2658	50.0000	
5	100PPBCN	0.098	99.4626	100.0000	
6	250PPBCN	0.245	254.6109	250.0000	
7	500PPBCN	0.474	497.7938	500.0000	