

LB94767


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

33387



A= 0.0000e+000

B= 3.0078e-004

C= 1.0458e-002

Rho= 0.9999424

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-6	0.000	-6				
0.05	0.050	0.059	0.009	161	0.0 %	161				
0.2	0.200	0.221	0.021	700	0.0 %	700				
2.5	2.500	2.514	0.014	8323	0.0 %	8323				
5	5.000	4.916	-0.084	16308	0.0 %	16308				
7.5	7.500	7.479	-0.021	24831	0.0 %	24831				
10	10.000	10.053	0.053	33387	0.0 %	33387				

LB94767 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Apr 2018 13:20:47

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000245.1	PPB	20 Apr 2018 13:24:05	S	Std
0.05 - 1	50.05	161	-	0.0500245.1	PPB	20 Apr 2018 13:26:13	S	Std
0.2 - 1	50.2	700	-	0.2000245.1	PPB	20 Apr 2018 13:28:21	S	Std
2.5 - 1	52.5	8323	-	2.5000245.1	PPB	20 Apr 2018 13:32:11	S	Std
5 - 1	55	16308	-	5.0000245.1	PPB	20 Apr 2018 13:34:18	S	Std
7.5 - 1	57.5	24831	-	7.5000245.1	PPB	20 Apr 2018 13:36:25	S	Std
10 - 1	510	33387	-	10.0000245.1	PPB	20 Apr 2018 13:38:33	S	Std
ICV07 - 1	ICV07	13842	4.1739	-245.1	PPB	20 Apr 2018 13:41:07	U	SMPL
ICB07 - 1	ICB07	-122	-0.0262	-245.1	PPB	20 Apr 2018 13:43:13	U	SMPL
CCV89 - 1	CCV89	17212	5.1875	-245.1	PPB	20 Apr 2018 13:45:20	U	SMPL
CCB89 - 1	CCB89	-151	-0.0350	-245.1	PPB	20 Apr 2018 13:47:27	U	SMPL
CRA - 1	CRA	146	0.0544	-245.1	PPB	20 Apr 2018 13:52:15	U	SMPL
HighStd - 1	HighStd	33475	10.0792	-245.1	PPB	20 Apr 2018 13:54:21	U	SMPL
ChkStd - 1	ChkStd	23548	7.0933	-245.1	PPB	20 Apr 2018 13:56:28	U	SMPL
PB108424BL - 1	PBW	-232	-0.0593	-245.1	PPB	20 Apr 2018 13:58:37	U	SMPL
PB108424BS - 1	LCSW	12845	3.8740	-245.1	PPB	20 Apr 2018 14:00:46	U	SMPL
J2473-01 - 1	041818-DBRI-F-D-B	-143	-0.0326	-245.1	PPB	20 Apr 2018 14:02:54	U	SMPL
J2473-02 - 1	041818-DBRI-F-D-S	-38	-0.0010	-245.1	PPB	20 Apr 2018 14:05:02	U	SMPL
J2473-03 - 1	041818-DBRI-F-U-B	-22	0.0038	-245.1	PPB	20 Apr 2018 14:07:11	U	SMPL
J2473-04 - 1	041818-DBRI-F-U-S	-19	0.0047	-245.1	PPB	20 Apr 2018 14:09:19	U	SMPL
J2473-05 - 1	041818-SIDI-F-D-B	-15	0.0059	-245.1	PPB	20 Apr 2018 14:11:26	U	SMPL
CCV90 - 1	CCV90	16975	5.1163	-245.1	PPB	20 Apr 2018 14:13:33	U	SMPL
CCB90 - 1	CCB90	-145	-0.0332	-245.1	PPB	20 Apr 2018 14:15:40	U	SMPL
J2473-06 - 1	041818-SIDI-F-D-M	-24	0.0032	-245.1	PPB	20 Apr 2018 14:17:47	U	SMPL
J2473-07 - 1	041818-SIDI-F-D-S	-20	0.0044	-245.1	PPB	20 Apr 2018 14:19:55	U	SMPL
J2473-08 - 1	041818-SIDI-F-U-B	-15	0.0059	-245.1	PPB	20 Apr 2018 14:22:01	U	SMPL
J2473-09 - 1	041818-SIDI-F-U-M	-22	0.0038	-245.1	PPB	20 Apr 2018 14:24:09	U	SMPL
J2473-10 - 1	041818-SIDI-F-U-S	-19	0.0047	-245.1	PPB	20 Apr 2018 14:26:16	U	SMPL
J2473-11 - 1	041818-SIDW-F-D-B	20	0.0165	-245.1	PPB	20 Apr 2018 14:28:23	U	SMPL
J2473-12 - 1	041818-SIDW-F-D-M	-18	0.0050	-245.1	PPB	20 Apr 2018 14:30:31	U	SMPL
J2473-13 - 1	041818-SIDW-F-D-S	-9	0.0078	-245.1	PPB	20 Apr 2018 14:32:40	U	SMPL
J2473-14 - 1	041818-SIDW-F-U-B	-6	0.0087	-245.1	PPB	20 Apr 2018 14:34:48	U	SMPL
J2473-15 - 1	041818-SIDW-F-U-M	-7	0.0084	-245.1	PPB	20 Apr 2018 14:36:55	U	SMPL
CCV91 - 1	CCV91	17079	5.1475	-245.1	PPB	20 Apr 2018 14:39:02	U	SMPL
CCB91 - 1	CCB91	-132	-0.0292	-245.1	PPB	20 Apr 2018 14:41:09	U	SMPL
J2473-16 - 1	041818-SIDW-F-U-S	-19	0.0047	-245.1	PPB	20 Apr 2018 14:43:17	U	SMPL
J2473-16DUP - 1	041818-SIDW-F-U-SDUP	-5	0.0090	-245.1	PPB	20 Apr 2018 14:45:24	U	SMPL
J2473-16MS - 1	041818-SIDW-F-U-SMS	14968	4.5126	-245.1	PPB	20 Apr 2018 14:47:31	U	SMPL
J2473-16MSD - 1	041818-SIDW-F-U-SMSD	15087	4.5484	-245.1	PPB	20 Apr 2018 14:49:38	U	SMPL
J2473-16LX5 - 1		-127	-0.0277	-245.1	PPB	20 Apr 2018 14:51:46	U	SMPL
J2473-16A - 1		15111	4.5556	-245.1	PPB	20 Apr 2018 14:53:54	U	SMPL
CCV92 - 1	CCV92	17153	5.1698	-245.1	PPB	20 Apr 2018 14:56:02	U	SMPL
CCB92 - 1	CCB92	-145	-0.0332	-245.1	PPB	20 Apr 2018 14:58:09	U	SMPL