


ISM02.4

LB94783
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40421



A= 0.0000e+000

B= 2.4784e-004

C= -2.3229e-002

Rho= 0.9999578

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.001	0.001	97	0.000	97					-
0.2	0.200	0.219	0.019	978	0.0 %	978					9
2.5	2.500	2.502	0.002	10189	0.0 %	10189					0
5.0	5.000	4.935	-0.065	20006	0.0 %	20006					-1
7.5	7.500	7.548	0.048	30548	0.0 %	30548					1
10.0	10.000	9.995	-0.005	40421	0.0 %	40421					0

NB
4/20/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	97	-		0 ISM02.4	PPB	4/20/2018 15:18	Std
	0.2 S0.2	978	-		0.2 ISM02.4	PPB	4/20/2018 15:20	Std
	2.5 S2.5	10189	-		2.5 ISM02.4	PPB	4/20/2018 15:22	Std
	5 S5	20006	-		5 ISM02.4	PPB	4/20/2018 15:24	Std
	7.5 S7.5	30548	-		7.5 ISM02.4	PPB	4/20/2018 15:27	Std
	10 S10	40421	-		10 ISM02.4	PPB	4/20/2018 15:29	Std
ICV	ICV	15803	3.8935 -		ISM02.4	PPB	4/20/2018 15:32	SMPL
ICB	ICB	-62	-0.0386 -		ISM02.4	PPB	4/20/2018 15:34	SMPL
CCV093	CCV093	20179	4.978 -		ISM02.4	PPB	4/20/2018 15:36	SMPL
CCB093	CCB093	-55	-0.0369 -		ISM02.4	PPB	4/20/2018 15:38	SMPL
PB108443BL	PBW010	-24	-0.0292 -		ISM02.4	PPB	4/20/2018 15:41	SMPL
PB108443TB	LEB010	139	0.0112 -		ISM02.4	PPB	4/20/2018 15:43	SMPL
J2435-01	MH0S28	-20	-0.0282 -		ISM02.4	PPB	4/20/2018 15:45	SMPL
J2435-02	MH0S28D	45	-0.0121 -		ISM02.4	PPB	4/20/2018 15:48	SMPL
J2435-03	MH0S28S	4133	1.0011 -		ISM02.4	PPB	4/20/2018 15:50	SMPL
CCV094	CCV094	20122	4.9639 -		ISM02.4	PPB	4/20/2018 15:52	SMPL
CCB094	CCB094	-55	-0.0369 -		ISM02.4	PPB	4/20/2018 15:54	SMPL