


7471B

LB94789
INSTRUMENT ID: CV 2

Linear

 μ Abs.:

41107



A= 0.0000e+000

B= 2.4514e-004

C= -1.4652e-002

Rho= 0.9998460

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	-4	0.000	-4				
0.2	0.200	0.197	-0.003	865	0.0 %	865				
2.5	2.500	2.539	0.039	10419	0.0 %	10419				
5	5.000	5.048	0.048	20652	0.0 %	20652				
7.5	7.500	7.369	-0.131	30120	0.0 %	30120				
10	10.000	10.062	0.062	41107	0.0 %	41107				

LB94789 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Apr 2018 16:14:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-4	-	0.0000	7471B	PPB	20 Apr 2018 16:35:55	S	Std
0.2 - 1	50.2	865	-	0.2000	7471B	PPB	20 Apr 2018 16:41:55	S	Std
2.5 - 1	52.5	10419	-	2.5000	7471B	PPB	20 Apr 2018 16:44:13	S	Std
5 - 1	55	20652	-	5.0000	7471B	PPB	20 Apr 2018 16:46:30	S	Std
7.5 - 1	57.5	30120	-	7.5000	7471B	PPB	20 Apr 2018 16:48:47	S	Std
10 - 1	510	41107	-	10.0000	7471B	PPB	20 Apr 2018 16:53:39	S	Std
ICV08 - 1	ICV08	16117	3.9362	-7471B	PPB	20 Apr 2018 16:59:50	U	SMPL	
ICB08 - 1	ICB08	-152	-0.0519	-7471B	PPB	20 Apr 2018 17:02:06	U	SMPL	
CCV95 - 1	CCV95	20617	5.0393	-7471B	PPB	20 Apr 2018 17:04:22	U	SMPL	
CCB95 - 1	CCB95	-184	-0.0598	-7471B	PPB	20 Apr 2018 17:06:37	U	SMPL	
CRA - 1	CRA	840	0.1913	-7471B	PPB	20 Apr 2018 17:08:54	U	SMPL	
HighStd - 1	HighStd	41902	10.2570	-7471B	PPB	20 Apr 2018 17:11:10	U	SMPL	
ChkStd - 1	ChkStd	29135	7.1274	-7471B	PPB	20 Apr 2018 17:13:26	U	SMPL	
PB108436BL - 1	PBS	-233	-0.0718	-7471B	PPB	20 Apr 2018 17:15:44	U	SMPL	
PB108436BS - 1	LCSS	15027	3.6690	-7471B	PPB	20 Apr 2018 17:18:02	U	SMPL	
J2460-01 - 1	MDP-B-1-0-6	1627	0.3842	-7471B	PPB	20 Apr 2018 17:20:19	U	SMPL	
J2460-02 - 1	MDP-B-1-18-20	513	0.1111	-7471B	PPB	20 Apr 2018 17:22:37	U	SMPL	
J2460-03 - 1	MDP-B-2-0-6	5972	1.4493	-7471B	PPB	20 Apr 2018 17:24:54	U	SMPL	
J2460-04 - 1	MDP-B-2-18-20	376	0.0775	-7471B	PPB	20 Apr 2018 17:27:12	U	SMPL	
J2461-01 - 1	N-P001-S002-0001-01	2227	0.5313	-7471B	PPB	20 Apr 2018 17:29:27	U	SMPL	
CCV96 - 1	CCV96	20815	5.0878	-7471B	PPB	20 Apr 2018 17:31:43	U	SMPL	
CCB96 - 1	CCB96	-150	-0.0514	-7471B	PPB	20 Apr 2018 17:33:59	U	SMPL	
J2461-01DUP - 1	N-P001-S002-0001-01DUP	2454	0.5869	-7471B	PPB	20 Apr 2018 17:36:15	U	SMPL	
J2461-02 - 1	J2461-01MS	18148	4.4341	-7471B	PPB	20 Apr 2018 17:38:31	U	SMPL	
J2461-03 - 1	J2461-01MSD	18060	4.4125	-7471B	PPB	20 Apr 2018 17:40:47	U	SMPL	
J2461-04 - 1	N-P001-S002-0001-02	2017	0.4798	-7471B	PPB	20 Apr 2018 17:43:03	U	SMPL	
J2477-01 - 1	FK-PS-01-022	508	0.1099	-7471B	PPB	20 Apr 2018 17:45:20	U	SMPL	
J2480-01 - 1	EPE-118-1(10-12)	298	0.0584	-7471B	PPB	20 Apr 2018 17:47:37	U	SMPL	
J2480-02 - 1	EPE-118-1(15-17)	163	0.0253	-7471B	PPB	20 Apr 2018 17:49:54	U	SMPL	
J2480-03 - 1	EPE-L-7A(0-5)	8586	2.0901	-7471B	PPB	20 Apr 2018 17:52:12	U	SMPL	
J2461-01LX5 - 1		329	0.0660	-7471B	PPB	20 Apr 2018 17:54:30	U	SMPL	
J2461-01A - 1		18962	4.6336	-7471B	PPB	20 Apr 2018 17:56:46	U	SMPL	
CCV97 - 1	CCV97	20873	5.1021	-7471B	PPB	20 Apr 2018 17:59:03	U	SMPL	
CCB97 - 1	CCB97	-136	-0.0480	-7471B	PPB	20 Apr 2018 18:01:19	U	SMPL	