



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94795

Slope : 99.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	04/20/2018	17:05
2	CAL2	1	Water			20.4	7.00	04/20/2018	17:07
3	CAL3	1	Water			20.4	10.03	04/20/2018	17:09
4	ICV	1	Water			20.5	7.00	04/20/2018	17:11
5	CCV1	1	Water			20.3	2.00	04/20/2018	17:13
6	J2498-01	1	Water			19.2	9.22	04/20/2018	17:15
7	J2498-01DUP	1	Water			19.2	9.23	04/20/2018	17:17
8	J2501-02	1	Water			17.0	7.32	04/20/2018	17:19
9	CCV2	1	Water			20.5	12.00	04/20/2018	17:21

WORKLIST(Hardcopy Internal Chain)

LB 94795

WorkList Name : PH J2501

WorkList ID : 111074

Date : 4/20/2018 4:53:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/20/2018	Water	J2498-01	pH	Cool 4 deg C	PSEG01	G41	VB16206-VB16122	04/20/2018	9040C
04/20/2018	Water	J2501-02	pH	Cool 4 deg C	PSEG01	G61	FADW4807L	04/20/2018	9040C

Date/Time 04/20/18 5:00

Received by: SP

Relinquished by: CP

Date/Time 04/20/18 5:45

Received by: CP

Relinquished by: SP