

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/23/2018

TEMP1 IN: 103 04/23/2018 07:25 TEMP OUT: 103 04/23/2018 08:30

TEMP2 IN: 104 04/23/2018 09:20 TEMP OUT: 104 04/23/2018 10:30

TEMP3 IN: 105 04/23/2018 12:30 TEMP OUT: 105 04/23/2018 14:10

TEMP4 IN: 105 04/23/2018 15:00 TEMP OUT: 105 04/23/2018 16:35

Run Number: LB94805

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB94805BL	LB94805BL	1.4068	1.4068	100	1.4068	1.4068	1.4068	0.0000	0
2	LB94805BS	LB94805BS	1.4056	1.4056	100	1.4560	1.4560	1.4560	0.0504	504
3	J2492-01	041918-DBRI-E-D-B	1.4089	1.4089	200	1.4247	1.4247	1.4247	0.0158	79
4	J2492-01DUP	041918-DBRI-E-D-BDUP	1.4262	1.4262	200	1.4417	1.4417	1.4417	0.0155	77.5
5	J2492-02	041918-DBRI-E-D-S	1.4198	1.4198	200	1.4390	1.4390	1.4390	0.0192	96
6	J2492-03	041918-DBRI-E-U-B	1.4060	1.4060	200	1.4219	1.4219	1.4219	0.0159	79.5
7	J2492-04	041918-DBRI-E-U-S	1.3969	1.3969	200	1.4071	1.4071	1.4071	0.0102	51
8	J2492-05	041918-JETB-E-D-M	1.4342	1.4342	200	1.4523	1.4523	1.4523	0.0181	90.5
9	J2492-06	041918-JETB-E-U-M	1.4015	1.4015	200	1.4240	1.4240	1.4240	0.0225	112.5
10	J2492-07	041918-SIDW-E-D-B	1.4085	1.4085	200	1.4233	1.4233	1.4233	0.0148	74
11	J2492-08	041918-SIDW-E-D-M	1.4108	1.4108	200	1.4244	1.4244	1.4244	0.0136	68
12	J2492-09	041918-SIDW-E-D-S	1.4112	1.4112	200	1.4211	1.4211	1.4211	0.0099	49.5
13	J2492-10	041918-SIDW-E-U-B	1.3985	1.3985	200	1.4105	1.4105	1.4105	0.0120	60
14	J2492-11	041918-SIDW-E-U-M	1.4241	1.4241	300	1.4381	1.4381	1.4381	0.0140	46.7
15	J2492-12	041918-SIDW-E-U-S	1.4073	1.4073	300	1.4191	1.4191	1.4191	0.0118	39.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 04/23/2018

TEMP1 IN: 103 04/23/2018 07:25 TEMP OUT: 103 04/23/2018 08:30

TEMP2 IN: 104 04/23/2018 09:20 TEMP OUT: 104 04/23/2018 10:30

TEMP3 IN: 105 04/23/2018 12:30 TEMP OUT: 105 04/23/2018 14:10

TEMP4 IN: 105 04/23/2018 15:00 TEMP OUT: 105 04/23/2018 16:35

Run Number: LB94805

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J2494-01	042018-JETB-E-D-M	1.4140	1.4140	200	1.4287	1.4287	1.4287	0.0147	73.5
17	J2494-01DUP	042018-JETB-E-D-MDUP	1.4098	1.4098	200	1.4240	1.4240	1.4240	0.0142	71
18	J2494-02	042018-JETB-E-U-M	1.4183	1.4183	200	1.4351	1.4351	1.4351	0.0168	84
19	J2494-03	042018-SIDI-E-D-B	1.4098	1.4098	200	1.4306	1.4306	1.4306	0.0208	104
20	J2494-04	042018-SIDI-E-D-M	1.4095	1.4095	200	1.4288	1.4288	1.4288	0.0193	96.5
21	J2494-05	042018-SIDI-E-D-S	1.4093	1.4093	200	1.4251	1.4251	1.4251	0.0158	79
22	J2494-06	042018-SIDI-E-U-B	1.4041	1.4041	200	1.4251	1.4251	1.4251	0.0210	105
23	J2494-07	042018-SIDI-E-U-M	1.4094	1.4094	200	1.4272	1.4272	1.4272	0.0178	89
24	J2494-08	042018-SIDI-E-U-S	1.4001	1.4001	300	1.4088	1.4088	1.4088	0.0087	29

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS042318

WorkList ID : 111104

Date : 4/23/2018 8:10:27 AM

1394805

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2018	Water	J2492-01 C	TSS	Cool 4 deg C	HDR103	G41	041918-DBRI-E-D-B	04/19/2018	SM2540 D
04/26/2018	Water	J2492-02 C	TSS	Cool 4 deg C	HDR103	G41	041918-DBRI-E-D-S	04/19/2018	SM2540 D
04/26/2018	Water	J2492-03 C	TSS	Cool 4 deg C	HDR103	G41	041918-DBRI-E-U-B	04/19/2018	SM2540 D
04/26/2018	Water	J2492-04 C	TSS	Cool 4 deg C	HDR103	G41	041918-DBRI-E-U-S	04/19/2018	SM2540 D
04/26/2018	Water	J2492-05 C	TSS	Cool 4 deg C	HDR103	G41	041918-JETB-E-D-M	04/19/2018	SM2540 D
04/26/2018	Water	J2492-06 C	TSS	Cool 4 deg C	HDR103	G41	041918-JETB-E-U-M	04/19/2018	SM2540 D
04/26/2018	Water	J2492-07 C	TSS	Cool 4 deg C	HDR103	G41	041918-SIDW-E-D-B	04/19/2018	SM2540 D
04/26/2018	Water	J2492-08 C	TSS	Cool 4 deg C	HDR103	G41	041918-SIDW-E-D-M	04/19/2018	SM2540 D
04/26/2018	Water	J2492-09 C	TSS	Cool 4 deg C	HDR103	G41	041918-SIDW-E-D-S	04/19/2018	SM2540 D
04/26/2018	Water	J2492-10 C	TSS	Cool 4 deg C	HDR103	G41	041918-SIDW-E-U-B	04/19/2018	SM2540 D
04/26/2018	Water	J2492-11 C	TSS	Cool 4 deg C	HDR103	G41	041918-SIDW-E-U-M	04/19/2018	SM2540 D
04/26/2018	Water	J2492-12 C	TSS	Cool 4 deg C	HDR103	G41	041918-SIDW-E-U-S	04/19/2018	SM2540 D
04/27/2018	Water	J2494-01 C	TSS	Cool 4 deg C	HDR103	G51	042018-JETB-E-D-M	04/20/2018	SM2540 D
04/27/2018	Water	J2494-02 C	TSS	Cool 4 deg C	HDR103	G51	042018-JETB-E-U-M	04/20/2018	SM2540 D
04/27/2018	Water	J2494-03 C	TSS	Cool 4 deg C	HDR103	G51	042018-SIDI-E-D-B	04/20/2018	SM2540 D
04/27/2018	Water	J2494-04 C	TSS	Cool 4 deg C	HDR103	G51	042018-SIDI-E-D-M	04/20/2018	SM2540 D
04/27/2018	Water	J2494-05 C	TSS	Cool 4 deg C	HDR103	G51	042018-SIDI-E-D-S	04/20/2018	SM2540 D
04/27/2018	Water	J2494-06 C	TSS	Cool 4 deg C	HDR103	G51	042018-SIDI-E-U-B	04/20/2018	SM2540 D
04/27/2018	Water	J2494-07 C	TSS	Cool 4 deg C	HDR103	G51	042018-SIDI-E-U-M	04/20/2018	SM2540 D
04/27/2018	Water	J2494-08 C	TSS	Cool 4 deg C	HDR103	G51	042018-SIDI-E-U-S	04/20/2018	SM2540 D

Date/Time 04/23/18 09:00
 Received by: 
 Relinquished by: 

Date/Time 04/23/18 13:00
 Received by: 
 Relinquished by: 