

7471B


LB94809

INSTRUMENT ID: CY1

Linear

 μ Abs.:

4180E



A= 0.0000e+000

B= 2.3550e-004

C= -1.6329e-002

Rho= 0.9990348

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	97	0.000	97				
0.2	0.200	0.223	0.023	1016	0.0 %	1016				
2.5	2.500	2.483	-0.017	10613	0.0 %	10613				
5	5.000	4.839	-0.161	20619	0.0 %	20619				
7.5	7.500	7.819	0.319	33271	0.0 %	33271				
10	10.000	9.829	-0.171	41806	0.0 %	41806				

LB94809 INSTRUMENT : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 23 Apr 2018 11:00:05				
Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	97	-	0.00007471B	PPB	23 Apr 2018 11:05:57	S	Std
0.2 - 1	50.2	1016	-	0.20007471B	PPB	23 Apr 2018 11:08:15	S	Std
2.5 - 1	52.5	10613	-	2.50007471B	PPB	23 Apr 2018 11:10:33	S	Std
5 - 1	55	20619	-	5.00007471B	PPB	23 Apr 2018 11:12:50	S	Std
7.5 - 1	57.5	33271	-	7.50007471B	PPB	23 Apr 2018 11:15:07	S	Std
10 - 1	57.5	41806	-	10.00007471B	PPB	23 Apr 2018 11:22:36	S	Std
ICV09 - 1	ICV09	17097	4.0100	-7471B	PPB	23 Apr 2018 11:25:26	U	SMPL
ICB09 - 1	ICB09	-97	-0.0392	-7471B	PPB	23 Apr 2018 11:27:41	U	SMPL
CCV01 - 1	CCV01	20803	4.8828	-7471B	PPB	23 Apr 2018 11:29:57	U	SMPL
CCB01 - 1	CCB01	-98	-0.0394	-7471B	PPB	23 Apr 2018 11:32:13	U	SMPL
CRA - 1	CRA	989	0.2166	-7471B	PPB	23 Apr 2018 11:34:29	U	SMPL
HighStd - 1	HighStd	42633	10.0238	-7471B	PPB	23 Apr 2018 11:36:45	U	SMPL
ChkStd - 1	ChkStd	29898	7.0247	-7471B	PPB	23 Apr 2018 11:39:01	U	SMPL
PB108459BL - 1	PBS	-212	-0.0663	-7471B	PPB	23 Apr 2018 11:41:20	U	SMPL
PB108459BS - 1	LCSS	17686	4.1487	-7471B	PPB	23 Apr 2018 11:43:39	U	SMPL
J2152-03 - 1	BU-03-042018	1757	0.3974	-7471B	PPB	23 Apr 2018 11:45:56	U	SMPL
J2161-05 - 1	OR-04-042018	720	0.1532	-7471B	PPB	23 Apr 2018 11:48:14	U	SMPL
J2482-02 - 1	NORTHLAWN-PRECHAR-AREA-1	11198	2.6208	-7471B	PPB	23 Apr 2018 11:50:32	U	SMPL
J2482-04 - 1	NORTHLAWN-PRECHAR-AREA-2	18518	4.3447	-7471B	PPB	23 Apr 2018 11:52:49	U	SMPL
J2482-06 - 1	NORTHLAWN-PRECHAR-AREA-3	36607	8.6047	-7471B	PPB	23 Apr 2018 11:55:05	U	SMPL
CCV02 - 1	CCV02	20397	4.7872	-7471B	PPB	23 Apr 2018 11:57:22	U	SMPL
CCB02 - 1	CCB02	-85	-0.0363	-7471B	PPB	23 Apr 2018 11:59:40	U	SMPL
J2490-01 - 1	MDP-B-3-0-6	5881	1.3687	-7471B	PPB	23 Apr 2018 12:01:57	U	SMPL
J2490-02 - 1	MDP-B-3-18-20	145	0.0178	-7471B	PPB	23 Apr 2018 12:04:13	U	SMPL
J2493-01 - 1	RS-RO-RT-27438	3341	0.7705	-7471B	PPB	23 Apr 2018 12:06:29	U	SMPL
J2495-01 - 1	RS-RO-1-7-A	2197	0.5011	-7471B	PPB	23 Apr 2018 12:08:45	U	SMPL
J2495-03 - 1	RS-RO-1-7-B	2868	0.6591	-7471B	PPB	23 Apr 2018 12:11:02	U	SMPL
J2496-01 - 1	RS-RO-1-3	3677	0.8496	-7471B	PPB	23 Apr 2018 12:13:19	U	SMPL
J2499-01 - 1	355-285-279	6042	1.4066	-7471B	PPB	23 Apr 2018 12:15:36	U	SMPL
J2499-03 - 1	278	3356	0.7740	-7471B	PPB	23 Apr 2018 12:17:54	U	SMPL
J2499-03DUP - 1	278DUP	3437	0.7931	-7471B	PPB	23 Apr 2018 12:20:12	U	SMPL
J2499-03MS - 1	278MS	19284	4.5251	-7471B	PPB	23 Apr 2018 12:22:28	U	SMPL
CCV03 - 1	CCV03	20665	4.8503	-7471B	PPB	23 Apr 2018 12:24:44	U	SMPL
CCB03 - 1	CCB03	-82	-0.0356	-7471B	PPB	23 Apr 2018 12:27:00	U	SMPL
J2499-03MSD - 1	278MSD	20004	4.6946	-7471B	PPB	23 Apr 2018 12:29:16	U	SMPL
J2500-01 - 1	RBR-400028-2171	6498	1.5140	-7471B	PPB	23 Apr 2018 12:31:33	U	SMPL
J2503-01 - 1	DRUM-2-1-PPE	-39	-0.0255	-7471B	PPB	23 Apr 2018 12:33:50	U	SMPL
J2503-03 - 1	CONCRETE	4734	1.0985	-7471B	PPB	23 Apr 2018 12:36:06	U	SMPL
J2503-05 - 1	RBR-47184RT	98431	23.1643	-7471B	PPB	23 Apr 2018 12:38:23	U	SMPL
J2503-07 - 1	CHRT25773	74998	17.6458	-7471B	PPB	23 Apr 2018 12:40:40	U	SMPL
J2499-03A - 1		20769	4.8748	-7471B	PPB	23 Apr 2018 12:45:19	U	SMPL
J2499-03LX5 - 1		665	0.1403	-7471B	PPB	23 Apr 2018 12:48:09	U	SMPL
J2503-05DLX5 - 1		19855	4.6595	-7471B	PPB	23 Apr 2018 12:50:47	U	SMPL
J2503-07DLX5 - 1		14662	3.4366	-7471B	PPB	23 Apr 2018 12:53:03	U	SMPL
CCV04 - 1	CCV04	20502	4.8119	-7471B	PPB	23 Apr 2018 12:55:19	U	SMPL
CCB04 - 1	CCB04	-97	-0.0392	-7471B	PPB	23 Apr 2018 12:57:36	U	SMPL