



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94812

Ambient Barometric Pressure (mmHg): 771

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	78.8	04/23/2018	09:02
2	J2498-02		1	>100.0	>212.0	>212.0	04/23/2018	09:40
3	J2498-02DUP		1	>100.0	>212.0	>212.0	04/23/2018	10:10
4	J2501-03		1	>100.0	>212.0	>212.0	04/23/2018	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB 94872

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-042318

WorkList ID : 111109

Date : 4/23/2018 8:12:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/25/2018	Water	J2498-02	C Flash Point	Cool 4 deg C	PSEG01	G41	VB16206-VB16122	04/20/2018	1010A
04/25/2018	Water	J2501-03	A Flash Point	Cool 4 deg C	PSEG01	G61	FADW4807L	04/20/2018	1010A

Date/Time 04/23/18 8:20 AM
Received by: KS
Relinquished by: CP

Date/Time 04/23/18 10:50
Received by: CP
Relinquished by: KS