

=====

Test results

Aquakem 7.2AQ1

21394813
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/23/2018 11:59

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.919	0.0	0.093	
ICB1	-0.002	0.0	0.004	
CCV1	1.008	0.0	0.102	
CCB1	-0.003	0.0	0.004	
PB108450BL	-0.002	0.0	0.004	
PB108450BS	0.907	0.0	0.092	
J2471-01	0.005	0.0	0.005	
J2471-01DUP	-0.003	0.0	0.004	
J2471-01MS	0.972	0.0	0.099	
J2471-01MSD	0.953	0.0	0.097	
CCV2	0.975	0.0	0.099	
CCB2	-0.003	0.0	0.004	

N	12
Mean	0.477
SD	0.5004
CV%	104.85

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 23 10:20:16 2018

Mon Apr 23 11:53:16 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0512	mg/l	4/23/2018 10:20:16	
0.1PPM	A	Phenolics-	P	0.1004	mg/l	4/23/2018 10:20:17	
0.25PPM	A	Phenolics-	P	0.2504	mg/l	4/23/2018 10:20:18	
0.50PPM	A	Phenolics-	P	0.4989	mg/l	4/23/2018 10:20:19	
1.0PPM	A	Phenolics-	P	0.9979	mg/l	4/23/2018 10:20:20	
2.0PPM	A	Phenolics-	P	2.0012	mg/l	4/23/2018 10:20:21	
ICV1	S	Phenolics-	P	0.9189	mg/l	4/23/2018 11:50:28	
ICB1	S	Phenolics-	P	-0.0019	mg/l	4/23/2018 11:50:29	
CCV1	S	Phenolics-	P	1.0078	mg/l	4/23/2018 11:50:30	
CCB1	S	Phenolics-	P	-0.0025	mg/l	4/23/2018 11:50:31	
PB108450BL	S	Phenolics-	P	-0.0019	mg/l	4/23/2018 11:50:32	
PB108450BS	S	Phenolics-	P	0.9075	mg/l	4/23/2018 11:50:33	
J2471-01	S	Phenolics-	P	0.0054	mg/l	4/23/2018 11:50:34	
J2471-01DUP	S	Phenolics-	P	-0.0033	mg/l	4/23/2018 11:50:35	
J2471-01MS	S	Phenolics-	P	0.972	mg/l	4/23/2018 11:50:36	
J2471-01MSD	S	Phenolics-	P	0.9526	mg/l	4/23/2018 11:50:37	
CCV2	S	Phenolics-	P	0.9751	mg/l	4/23/2018 11:50:38	
CCB2	S	Phenolics-	P	-0.0031	mg/l	4/23/2018 11:50:39	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/23/2018 10:48


Test Phenolics-

Accepted 4/23/2018 10:48

Factor 10.33
Bias 0.004

Coeff. of det. 0.999997

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0512	0.0500	
2	Phenol-2	0.014	0.1004	0.1000	
3	Phenol-2	0.029	0.2504	0.2500	
4	Phenol-2	0.053	0.4989	0.5000	
5	Phenol-2	0.101	0.9979	1.0000	
6	Phenol-2	0.198	2.0012	2.0000	