


245.1

LB94815
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33570



A= 0.0000e+000

B= 2.9867e-004

C= -2.5952e-002

Rho= 0.9999403

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	37	0.000	37				
0.05	0.050	0.051	0.001	256	0.0 %	256				
0.2	0.200	0.198	-0.002	751	0.0 %	751				
2.5	2.500	2.482	-0.018	8397	0.0 %	8397				
5	5.000	5.089	0.089	17126	0.0 %	17126				
7.5	7.500	7.445	-0.055	25014	0.0 %	25014				
10	10.000	10.000	0.000	33570	0.0 %	33570				

LB94815 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 23 Apr 2018 14:54:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	37	-	0.0000	245.1	PPB	23 Apr 2018 14:57:34	S	Std
0.05 - 1	50.05	256	-	0.0500	245.1	PPB	23 Apr 2018 14:59:42	S	Std
0.2 - 1	50.2	751	-	0.2000	245.1	PPB	23 Apr 2018 15:01:51	S	Std
2.5 - 1	52.5	8397	-	2.5000	245.1	PPB	23 Apr 2018 15:03:59	S	Std
5 - 1	55	17126	-	5.0000	245.1	PPB	23 Apr 2018 15:06:07	S	Std
7.5 - 1	57.5	25014	-	7.5000	245.1	PPB	23 Apr 2018 15:08:14	S	Std
10 - 1	510	33570	-	10.0000	245.1	PPB	23 Apr 2018 15:10:24	S	Std
ICV10 - 1	ICV10	13990	4.1524	-245.1	PPB	23 Apr 2018 15:13:00	U	SMPL	
ICB10 - 1	ICB10	-96	-0.0546	-245.1	PPB	23 Apr 2018 15:15:06	U	SMPL	
CCV05 - 1	CCV05	17442	5.1834	-245.1	PPB	23 Apr 2018 15:17:14	U	SMPL	
CCB05 - 1	CCB05	-133	-0.0657	-245.1	PPB	23 Apr 2018 15:19:21	U	SMPL	
CRA - 1	CRA	225	0.0412	-245.1	PPB	23 Apr 2018 15:21:28	U	SMPL	
HighStd - 1	HighStd	34891	10.3948	-245.1	PPB	23 Apr 2018 15:23:35	U	SMPL	
ChkStd - 1	ChkStd	23826	7.0901	-245.1	PPB	23 Apr 2018 15:25:41	U	SMPL	
PB108491BL - 1	PBW	-207	-0.0878	-245.1	PPB	23 Apr 2018 15:27:51	U	SMPL	
PB108491BS - 1	LCSW	13688	4.0622	-245.1	PPB	23 Apr 2018 15:30:00	U	SMPL	
J2492-01 - 1	041918-DBRI-E-D-B	-148	-0.0702	-245.1	PPB	23 Apr 2018 15:32:08	U	SMPL	
J2492-02 - 1	041918-DBRI-E-D-S	-34	-0.0361	-245.1	PPB	23 Apr 2018 15:34:16	U	SMPL	
J2492-03 - 1	041918-DBRI-E-U-B	-13	-0.0298	-245.1	PPB	23 Apr 2018 15:36:25	U	SMPL	
J2492-04 - 1	041918-DBRI-E-U-S	-6	-0.0277	-245.1	PPB	23 Apr 2018 15:38:33	U	SMPL	
J2492-05 - 1	041918-JETB-E-D-M	-1	-0.0263	-245.1	PPB	23 Apr 2018 15:40:40	U	SMPL	
CCV06 - 1	CCV06	17323	5.1478	-245.1	PPB	23 Apr 2018 15:42:46	U	SMPL	
CCB06 - 1	CCB06	-127	-0.0639	-245.1	PPB	23 Apr 2018 15:44:53	U	SMPL	
J2492-06 - 1	041918-JETB-E-U-M	4	-0.0248	-245.1	PPB	23 Apr 2018 15:47:01	U	SMPL	
J2492-07 - 1	041918-SIDW-E-D-B	4	-0.0248	-245.1	PPB	23 Apr 2018 15:49:08	U	SMPL	
J2492-08 - 1	041918-SIDW-E-D-M	-22	-0.0325	-245.1	PPB	23 Apr 2018 15:51:14	U	SMPL	
J2492-09 - 1	041918-SIDW-E-D-S	-18	-0.0313	-245.1	PPB	23 Apr 2018 15:53:21	U	SMPL	
J2492-10 - 1	041918-SIDW-E-U-B	-29	-0.0346	-245.1	PPB	23 Apr 2018 15:55:29	U	SMPL	
J2492-11 - 1	041918-SIDW-E-U-M	-9	-0.0286	-245.1	PPB	23 Apr 2018 15:57:36	U	SMPL	
J2492-12 - 1	041918-SIDW-E-U-S	-17	-0.0310	-245.1	PPB	23 Apr 2018 15:59:44	U	SMPL	
J2494-01 - 1	042018-JETB-E-D-M	-37	-0.0370	-245.1	PPB	23 Apr 2018 16:01:52	U	SMPL	
J2494-02 - 1	042018-JETB-E-U-M	-16	-0.0307	-245.1	PPB	23 Apr 2018 16:04:01	U	SMPL	
J2494-03 - 1	042018-SIDI-E-D-B	-10	-0.0289	-245.1	PPB	23 Apr 2018 16:06:07	U	SMPL	
CCV07 - 1	CCV07	17155	5.0977	-245.1	PPB	23 Apr 2018 16:08:15	U	SMPL	
CCB07 - 1	CCB07	-136	-0.0666	-245.1	PPB	23 Apr 2018 16:10:21	U	SMPL	
J2494-04 - 1	042018-SIDI-E-D-M	-13	-0.0298	-245.1	PPB	23 Apr 2018 16:12:30	U	SMPL	
J2494-05 - 1	042018-SIDI-E-D-S	-24	-0.0331	-245.1	PPB	23 Apr 2018 16:14:37	U	SMPL	
J2494-06 - 1	042018-SIDI-E-U-B	-11	-0.0292	-245.1	PPB	23 Apr 2018 16:16:44	U	SMPL	
J2494-07 - 1	042018-SIDI-E-U-M	1	-0.0257	-245.1	PPB	23 Apr 2018 16:18:51	U	SMPL	
J2494-08 - 1	042018-SIDI-E-U-S	-9	-0.0286	-245.1	PPB	23 Apr 2018 16:20:59	U	SMPL	
J2494-08DUP - 1	042018-SIDI-E-U-SDUP	-18	-0.0313	-245.1	PPB	23 Apr 2018 16:23:06	U	SMPL	
J2494-08MS - 1	042018-SIDI-E-U-SMS	14120	4.1912	-245.1	PPB	23 Apr 2018 16:25:14	U	SMPL	
J2494-08MSD - 1	042018-SIDI-E-U-SMSD	13739	4.0774	-245.1	PPB	23 Apr 2018 16:27:23	U	SMPL	
J2494-08LX5 - 1		-101	-0.0561	-245.1	PPB	23 Apr 2018 16:29:31	U	SMPL	
J2494-08A - 1		13751	4.0810	-245.1	PPB	23 Apr 2018 16:31:38	U	SMPL	
CCV08 - 1	CCV08	17255	5.1275	-245.1	PPB	23 Apr 2018 16:33:45	U	SMPL	
CCB08 - 1	CCB08	-140	-0.0678	-245.1	PPB	23 Apr 2018 16:35:52	U	SMPL	