

LB 94835


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37286



A= 0.0000e+000

B= 2.6802e-004

C= 1.5980e-002

Rho= 0.9997174

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	-19	0.000	-19				
0.2	0.200	0.215	0.015	742	0.0 %	742				
2.5	2.500	2.569	0.069	9524	0.0 %	9524				
5	5.000	4.815	-0.185	17904	0.0 %	17904				
7.5	7.500	7.582	0.082	28228	0.0 %	28228				
10	10.000	10.009	0.009	37286	0.0 %	37286				

LB94835 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Apr 2018 10:04:59

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0007470A	PPB	24 Apr 2018 10:17:59	S	Std
0.2 - 1	90.2	742	-	0.20007470A	PPB	24 Apr 2018 10:23:45	S	Std
2.5 - 1	92.5	9524	-	2.50007470A	PPB	24 Apr 2018 10:25:53	S	Std
5 - 1	55	17904	-	5.00007470A	PPB	24 Apr 2018 10:28:01	S	Std
7.5 - 1	57.5	28228	-	7.50007470A	PPB	24 Apr 2018 10:30:09	S	Std
10 - 1	510	37286	-	10.00007470A	PPB	24 Apr 2018 10:32:18	S	Std
ICV11 - 1	ICV11	13384	3.6032	-7470A	PPB	24 Apr 2018 10:35:01	U	SMPL
ICB11 - 1	ICB11	-161	-0.0272	-7470A	PPB	24 Apr 2018 10:37:08	U	SMPL
CCV09 - 1	CCV09	18718	5.0328	-7470A	PPB	24 Apr 2018 10:39:15	U	SMPL
CCB09 - 1	CCB09	-194	-0.0360	-7470A	PPB	24 Apr 2018 10:41:22	U	SMPL
CRA - 1	CRA	754	0.2181	-7470A	PPB	24 Apr 2018 10:43:30	U	SMPL
HighStd - 1	HighStd	37150	9.9729	-7470A	PPB	24 Apr 2018 10:45:36	U	SMPL
ChkStd - 1	ChkStd	24765	6.6535	-7470A	PPB	24 Apr 2018 10:47:43	U	SMPL
PB108511BL - 1	PBW	-230	-0.0457	-7470A	PPB	24 Apr 2018 10:49:52	U	SMPL
PB108511BS - 1	LCSW	13787	3.7112	-7470A	PPB	24 Apr 2018 10:52:00	U	SMPL
PB108511TB - 1	PB108453TB	-121	-0.0165	-7470A	PPB	24 Apr 2018 10:54:08	U	SMPL
J2152-04 - 1	BU-03-042018	-42	0.0047	-7470A	PPB	24 Apr 2018 10:56:16	U	SMPL
J2152-04DUP - 1	BU-03-042018DUP	-32	0.0074	-7470A	PPB	24 Apr 2018 10:58:25	U	SMPL
J2152-04MS - 1	BU-03-042018MS	12740	3.4308	-7470A	PPB	24 Apr 2018 11:00:33	U	SMPL
J2152-04MSD - 1	BU-03-042018MSD	13669	3.6795	-7470A	PPB	24 Apr 2018 11:02:40	U	SMPL
CCV10 - 1	CCV10	18367	4.9387	-7470A	PPB	24 Apr 2018 11:04:47	U	SMPL
CCB10 - 1	CCB10	-209	-0.0400	-7470A	PPB	24 Apr 2018 11:06:54	U	SMPL
J2161-06 - 1	OR-03-042018	-56	0.0010	-7470A	PPB	24 Apr 2018 11:09:02	U	SMPL
J2482-07 - 1	NORTHLAWN-PRECHAR-AREA-1	-9	0.0136	-7470A	PPB	24 Apr 2018 11:11:08	U	SMPL
J2482-08 - 1	NORTHLAWN-PRECHAR-AREA-2	-14	0.0122	-7470A	PPB	24 Apr 2018 11:13:15	U	SMPL
J2482-09 - 1	NORTHLAWN-PRECHAR-AREA-3	-14	0.0122	-7470A	PPB	24 Apr 2018 11:15:22	U	SMPL
J2490-03 - 1	MDP-WC-1	-17	0.0114	-7470A	PPB	24 Apr 2018 11:17:29	U	SMPL
J2493-03 - 1	RS-RO-RT-27438	1	0.0162	-7470A	PPB	24 Apr 2018 11:19:37	U	SMPL
J2495-02 - 1	RS-RO-1-7-A	-21	0.0104	-7470A	PPB	24 Apr 2018 11:21:47	U	SMPL
J2495-04 - 1	RS-RO-1-7-B	-12	0.0128	-7470A	PPB	24 Apr 2018 11:23:55	U	SMPL
J2496-03 - 1	RS-RO-1-3	-10	0.0133	-7470A	PPB	24 Apr 2018 11:26:03	U	SMPL
J2499-02 - 1	355-285-279	-3	0.0152	-7470A	PPB	24 Apr 2018 11:28:10	U	SMPL
CCV11 - 1	CCV11	18577	4.9950	-7470A	PPB	24 Apr 2018 11:30:17	U	SMPL
CCB11 - 1	CCB11	-157	-0.0261	-7470A	PPB	24 Apr 2018 11:32:24	U	SMPL
J2499-04 - 1	278	-37	0.0061	-7470A	PPB	24 Apr 2018 11:34:31	U	SMPL
J2500-02 - 1	RBR-400028-2171	-22	0.0101	-7470A	PPB	24 Apr 2018 11:36:38	U	SMPL
J2503-02 - 1	DRUM-2-1-PPE	-13	0.0125	-7470A	PPB	24 Apr 2018 11:38:45	U	SMPL
J2503-04 - 1	CONCRETE	-28	0.0085	-7470A	PPB	24 Apr 2018 11:40:52	U	SMPL
J2503-06 - 1	RBR-47184RT	-8	0.0138	-7470A	PPB	24 Apr 2018 11:43:00	U	SMPL
J2503-08 - 1	CHRT25773	-7	0.0141	-7470A	PPB	24 Apr 2018 11:45:07	U	SMPL
PB108514BL - 1	PBW	-25	0.0093	-7470A	PPB	24 Apr 2018 11:47:15	U	SMPL
PB108514BS - 1	LCSW	14973	4.0290	-7470A	PPB	24 Apr 2018 11:49:23	U	SMPL
PB108514TB - 1	PB108457TB	-125	-0.0175	-7470A	PPB	24 Apr 2018 11:51:32	U	SMPL
J2498-05 - 1	VB16206-VB16122	-15	0.0120	-7470A	PPB	24 Apr 2018 11:53:39	U	SMPL
CCV12 - 1	CCV12	18744	5.0398	-7470A	PPB	24 Apr 2018 11:55:46	U	SMPL
CCB12 - 1	CCB12	-162	-0.0274	-7470A	PPB	24 Apr 2018 11:57:54	U	SMPL
J2498-05DUP - 1	VB16206-VB16122DUP	-11	0.0130	-7470A	PPB	24 Apr 2018 12:00:02	U	SMPL
J2498-05MS - 1	VB16206-VB16122MS	14194	3.8203	-7470A	PPB	24 Apr 2018 12:02:09	U	SMPL
J2498-05MSD - 1	VB16206-VB16122MSD	14148	3.8079	-7470A	PPB	24 Apr 2018 12:04:16	U	SMPL
J2501-03 - 1	FADW4807L	-139	-0.0213	-7470A	PPB	24 Apr 2018 12:06:23	U	SMPL
PB108515BL - 1	PBW	-23	0.0098	-7470A	PPB	24 Apr 2018 12:08:30	U	SMPL
PB108515BS - 1	LCSW	14019	3.7734	-7470A	PPB	24 Apr 2018 12:10:38	U	SMPL
J2498-01 - 1	VB16206-VB16122	11549	3.1113	-7470A	PPB	24 Apr 2018 12:12:46	U	SMPL
J2498-01DUP - 1	VB16206-VB16122DUP	10806	2.9122	-7470A	PPB	24 Apr 2018 12:17:40	U	SMPL
J2498-01MS - 1	VB16206-VB16122MS	25772	6.9234	-7470A	PPB	24 Apr 2018 12:19:49	U	SMPL
CCV13 - 1	CCV13	17967	4.8315	-7470A	PPB	24 Apr 2018 12:21:56	U	SMPL
CCB13 - 1	CCB13	-136	-0.0205	-7470A	PPB	24 Apr 2018 12:24:04	U	SMPL
J2498-01MSD - 1	VB16206-VB16122MSD	26101	7.0116	-7470A	PPB	24 Apr 2018 12:26:12	U	SMPL
J2501-02 - 1	FADW4807L	330	0.1044	-7470A	PPB	24 Apr 2018 12:28:19	U	SMPL
J2510-02 - 1	FADW4710L	1203	0.3384	-7470A	PPB	24 Apr 2018 12:30:27	U	SMPL
J2152-04LX5 - 1		-12	0.0128	-7470A	PPB	24 Apr 2018 12:32:34	U	SMPL
J2152-04A - 1		12923	3.4796	-7470A	PPB	24 Apr 2018 12:34:42	U	SMPL
J2498-05LX5 - 1		-99	-0.0106	-7470A	PPB	24 Apr 2018 12:36:49	U	SMPL
J2498-05A - 1		12657	3.4083	-7470A	PPB	24 Apr 2018 12:38:57	U	SMPL
J2498-01LX5 - 1		1991	0.5496	-7470A	PPB	24 Apr 2018 12:41:06	U	SMPL
J2498-01A - 1		24683	6.6315	-7470A	PPB	24 Apr 2018 12:43:14	U	SMPL
CCV14 - 1	CCV14	16967	4.5635	-7470A	PPB	24 Apr 2018 12:45:23	U	SMPL
CCB14 - 1	CCB14	-146	-0.0232	-7470A	PPB	24 Apr 2018 12:47:32	U	SMPL