

=====

Test results

Aquakem 7.2AQ1

21394840
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS.

4/24/2018 11:35

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.350	0.0	0.098	
ICB1	-0.542	0.0	0.004	
CCV1	249.965	0.0	0.244	
CCB1	-0.272	0.0	0.004	
PB108518BL	-0.247	0.0	0.004	
PB108518BS	99.135	0.0	0.099	
J2490-01	0.229	0.0	0.004	
J2490-01DUP	0.212	0.0	0.004	
J2490-01MS	34.405	0.0	0.037	
J2490-01MSD	34.222	0.0	0.037	
J2490-02	-0.418	0.0	0.004	
CCV2	262.385	0.0	0.256	
CCB2	-0.329	0.0	0.004	

N	13
Mean	59.700
SD	94.2763
CV%	157.92

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

4/24/2018 8:38


Test Total CN

Accepted 4/24/2018 8:38

Factor 1041
Bias 0.004

Coeff. of det. 0.999963

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.6078	0.0000	
2	5.0PPBCN	0.008	4.2448	5.0000	
3	10PPBCN	0.013	9.5300	10.0000	
4	50PPBCN	0.053	50.8620	50.0000	
5	100PPBCN	0.100	99.9846	100.0000	
6	250PPBCN	0.246	252.1051	250.0000	
7	500PPBCN	0.483	498.8813	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 24 11:32:09 2018

Tue Apr 24 11:35:33 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.6078	µg/l	4/24/2018 8:36:26	
5.0PPBCN	A	Total CN	P	4.2448	µg/l	4/24/2018 8:36:27	
10PPBCN	A	Total CN	P	9.53	µg/l	4/24/2018 8:36:28	
50PPBCN	A	Total CN	P	50.862	µg/l	4/24/2018 8:36:29	
100PPBCN	A	Total CN	P	99.9846	µg/l	4/24/2018 8:36:30	
250PPBCN	A	Total CN	P	252.1051	µg/l	4/24/2018 8:36:31	
500PPBCN	A	Total CN	P	498.8813	µg/l	4/24/2018 8:36:32	
ICV1	S	Total CN	P	97.3504	µg/l	4/24/2018 11:32:09	
ICB1	S	Total CN	P	-0.5423	µg/l	4/24/2018 11:32:10	
CCV1	S	Total CN	P	249.9652	µg/l	4/24/2018 11:32:11	
CCB1	S	Total CN	P	-0.2721	µg/l	4/24/2018 11:32:12	
PB108518BL	S	Total CN	P	-0.2465	µg/l	4/24/2018 11:32:13	
PB108518BS	S	Total CN	P	99.1352	µg/l	4/24/2018 11:32:14	
J2490-01	S	Total CN	P	0.2293	µg/l	4/24/2018 11:32:15	
J2490-01DUP	S	Total CN	P	0.212	µg/l	4/24/2018 11:32:16	
J2490-01MS	S	Total CN	P	34.4046	µg/l	4/24/2018 11:32:17	
J2490-01MSD	S	Total CN	P	34.2215	µg/l	4/24/2018 11:32:18	
J2490-02	S	Total CN	P	-0.4175	µg/l	4/24/2018 11:32:19	
CCV2	S	Total CN	P	262.385	µg/l	4/24/2018 11:35:32	
CCB2	S	Total CN	P	-0.3291	µg/l	4/24/2018 11:35:33	