



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94842

Ambient Barometric Pressure (mmHg): 765

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	82.1	04/24/2018	12:02
2	J2518-05		1	20.0	68.0	67.7	04/24/2018	12:20
3	J2518-05DUP		1	20.0	68.0	67.7	04/24/2018	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06