



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94843

Ambient Barometric Pressure (mmHg): 765

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	80.8	04/24/2018	13:10
2	J2510-03		1	>100.0	>212.0	>212.0	04/24/2018	13:32
3	J2510-03DUP		1	>100.0	>212.0	>212.0	04/24/2018	13:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB 94843

WorkList Name : FP42318

WorkList ID : 111143

Date : 4/23/2018 4:36:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/26/2018	Water	J2510-03	C Flash Point	Cool 4 deg C	PSEG02	G62	FADW4710L	04/23/2018	1010A

Date/Time 04/24/18 11:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 04/24/18 14:05
 Received by: CP
 Relinquished by: KS