


7471B

LB94851
INSTRUMENT ID: CY1

Linear

 μ Abs.:

41981



A= 0.0000e+000

B= 2.3941e-004

C= -2.3352e-002

Rho= 0.9999256

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	218	0.000	218				
0.2	0.200	0.209	0.009	972	0.0 %	972				
2.5	2.500	2.509	0.009	10576	0.0 %	10576				
5	5.000	4.901	-0.099	20569	0.0 %	20569				
7.5	7.500	7.525	0.025	31530	0.0 %	31530				
10	10.000	10.027	0.027	41981	0.0 %	41981				

LB94851 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Apr 2018 15:05:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB21 - 1	CCB21	72	-0.0061		-7471B	PPB	24 Apr 2018 18:30:33	U	SMPL
J2520-13 - 1	TP-6_4.0-4.5	2237	0.5122		-7471B	PPB	24 Apr 2018 18:32:51	U	SMPL
J2520-14 - 1	TP-5_0.0-0.5	5104	1.1986		-7471B	PPB	24 Apr 2018 18:35:08	U	SMPL
J2520-15 - 1	TP-5_2.0-2.5	7180	1.6956		-7471B	PPB	24 Apr 2018 18:52:35	U	SMPL
J2520-16 - 1	TP-5_4.0-4.5	583	0.1162		-7471B	PPB	24 Apr 2018 18:54:52	U	SMPL
J2151-09LX5 - 1		471	0.0894		-7471B	PPB	24 Apr 2018 18:57:09	U	SMPL
J2151-09A - 1		18880	4.4966		-7471B	PPB	24 Apr 2018 18:59:27	U	SMPL
J2512-03LX5 - 1		279	0.0434		-7471B	PPB	24 Apr 2018 19:01:43	U	SMPL
J2512-03A - 1		17905	4.2632		-7471B	PPB	24 Apr 2018 19:04:01	U	SMPL
J2520-11LX5 - 1		1349	0.2996		-7471B	PPB	24 Apr 2018 19:06:19	U	SMPL
J2520-11A - 1		23073	5.5005		-7471B	PPB	24 Apr 2018 19:08:36	U	SMPL
CCV22 - 1	CCV22	21210	5.0544		-7471B	PPB	24 Apr 2018 19:10:53	U	SMPL
CCB22 - 1	CCB22	51	-0.0111		-7471B	PPB	24 Apr 2018 19:13:10	U	SMPL
J2520-09DLX5 - 1		21979	5.2385		-7471B	PPB	24 Apr 2018 19:15:28	U	SMPL
CCV23 - 1	CCV23	20957	4.9939		-7471B	PPB	24 Apr 2018 19:17:45	U	SMPL
CCB23 - 1	CCB23	-40	-0.0329		-7471B	PPB	24 Apr 2018 19:20:03	U	SMPL

LB94851 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Apr 2018 15:05:54

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	218	-	0.0000	7471B	PPB	24 Apr 2018 15:09:11	S	Std
0.2 - 1	50.2	972	-	0.2000	7471B	PPB	24 Apr 2018 15:11:28	S	Std
2.5 - 1	52.5	10576	-	2.5000	7471B	PPB	24 Apr 2018 15:13:46	S	Std
5 - 1	55	20569	-	5.0000	7471B	PPB	24 Apr 2018 15:16:03	S	Std
7.5 - 1	57.5	31530	-	7.5000	7471B	PPB	24 Apr 2018 15:18:20	S	Std
10 - 1	510	41981	-	10.0000	7471B	PPB	24 Apr 2018 15:26:07	S	Std
ICV12 - 1	ICV12	16362	3.8938	-	7471B	PPB	24 Apr 2018 15:28:51	U	SMPL
ICB12 - 1	ICB12	74	-0.0056	-	7471B	PPB	24 Apr 2018 15:31:07	U	SMPL
CCV15 - 1	CCV15	20467	4.8766	-	7471B	PPB	24 Apr 2018 15:33:23	U	SMPL
CCB15 - 1	CCB15	62	-0.0085	-	7471B	PPB	24 Apr 2018 15:35:39	U	SMPL
CRA - 1	CRA	931	0.1995	-	7471B	PPB	24 Apr 2018 15:37:55	U	SMPL
HighStd - 1	HighStd	43124	10.3008	-	7471B	PPB	24 Apr 2018 15:40:11	U	SMPL
ChkStd - 1	ChkStd	29183	6.9632	-	7471B	PPB	24 Apr 2018 15:42:27	U	SMPL
PB108525BL - 1	PBS	-108	-0.0492	-	7471B	PPB	24 Apr 2018 15:44:45	U	SMPL
PB108525BS - 1	LCSS	14681	3.4914	-	7471B	PPB	24 Apr 2018 15:47:03	U	SMPL
J2151-07 - 1	AU-05-042318-A	2204	0.5043	-	7471B	PPB	24 Apr 2018 15:49:20	U	SMPL
J2151-09 - 1	AU-05-042318-B	1698	0.3832	-	7471B	PPB	24 Apr 2018 15:51:38	U	SMPL
J2151-09DUP - 1	AU-05-042318-BDUP	1578	0.3544	-	7471B	PPB	24 Apr 2018 15:56:37	U	SMPL
J2151-09MS - 1	AU-05-042318-BMS	17780	4.2333	-	7471B	PPB	24 Apr 2018 16:04:03	U	SMPL
J2151-09MSD - 1	AU-05-042318-BMSD	17655	4.2034	-	7471B	PPB	24 Apr 2018 16:06:19	U	SMPL
CCV16 - 1	CCV16	20941	4.9900	-	7471B	PPB	24 Apr 2018 16:08:35	U	SMPL
CCB16 - 1	CCB16	29	-0.0164	-	7471B	PPB	24 Apr 2018 16:10:52	U	SMPL
J2151-11 - 1	AU-05-042318-C	2927	0.6774	-	7471B	PPB	24 Apr 2018 16:13:08	U	SMPL
J2151-13 - 1	AU-05-042318-D	1739	0.3930	-	7471B	PPB	24 Apr 2018 16:15:24	U	SMPL
J2504-01 - 1	KS-RO-238	789	0.1655	-	7471B	PPB	24 Apr 2018 16:17:40	U	SMPL
J2505-02 - 1	S-1-COMP	2694	0.6216	-	7471B	PPB	24 Apr 2018 16:19:57	U	SMPL
J2505-04 - 1	S-2-COMP	4762	1.1167	-	7471B	PPB	24 Apr 2018 16:22:13	U	SMPL
J2505-06 - 1	S-3-COMP	5523	1.2989	-	7471B	PPB	24 Apr 2018 16:24:30	U	SMPL
J2508-01 - 1	TP-1	8859	2.0975	-	7471B	PPB	24 Apr 2018 16:26:48	U	SMPL
J2508-05 - 1	TP-2	24652	5.8785	-	7471B	PPB	24 Apr 2018 16:29:05	U	SMPL
J2508-09 - 1	TP-3	25254	6.0226	-	7471B	PPB	24 Apr 2018 16:31:23	U	SMPL
J2508-13 - 1	TP-4	23458	5.5926	-	7471B	PPB	24 Apr 2018 16:33:40	U	SMPL
CCV17 - 1	CCV17	21078	5.0228	-	7471B	PPB	24 Apr 2018 16:35:58	U	SMPL
CCB17 - 1	CCB17	53	-0.0107	-	7471B	PPB	24 Apr 2018 16:38:15	U	SMPL
J2508-17 - 1	TP-5	21677	5.1662	-	7471B	PPB	24 Apr 2018 16:40:33	U	SMPL
J2508-21 - 1	TP-6	35260	8.4181	-	7471B	PPB	24 Apr 2018 16:42:49	U	SMPL
J2509-01 - 1	WS-DRUM	906	0.1935	-	7471B	PPB	24 Apr 2018 16:45:06	U	SMPL
J2511-01 - 1	FK-GF-02-017-A	2715	0.6266	-	7471B	PPB	24 Apr 2018 16:47:24	U	SMPL
J2511-03 - 1	FK-GF-02-017-B	4350	1.0181	-	7471B	PPB	24 Apr 2018 16:49:41	U	SMPL
J2515-01 - 1	S-1-A	12794	3.0396	-	7471B	PPB	24 Apr 2018 16:51:58	U	SMPL
J2515-03 - 1	S-2-A	7844	1.8545	-	7471B	PPB	24 Apr 2018 16:54:15	U	SMPL
J2515-05 - 1	S-3-A	729	0.1512	-	7471B	PPB	24 Apr 2018 16:56:33	U	SMPL
PB108526BL - 1	PBS	-2	-0.0238	-	7471B	PPB	24 Apr 2018 16:58:50	U	SMPL
PB108526BS - 1	LCSS	14551	3.4602	-	7471B	PPB	24 Apr 2018 17:01:07	U	SMPL
CCV18 - 1	CCV18	21192	5.0501	-	7471B	PPB	24 Apr 2018 17:03:24	U	SMPL
CCB18 - 1	CCB18	-9	-0.0255	-	7471B	PPB	24 Apr 2018 17:05:40	U	SMPL
J2512-01 - 1	TP-1-1-3	8906	2.1088	-	7471B	PPB	24 Apr 2018 17:07:57	U	SMPL
J2512-02 - 1	TP-2-12	990	0.2137	-	7471B	PPB	24 Apr 2018 17:10:13	U	SMPL
J2512-03 - 1	TP-3-2	1320	0.2927	-	7471B	PPB	24 Apr 2018 17:12:30	U	SMPL
J2512-03DUP - 1	TP-3-2DUP	1253	0.2766	-	7471B	PPB	24 Apr 2018 17:14:46	U	SMPL
J2512-03MS - 1	TP-3-2MS	17122	4.0757	-	7471B	PPB	24 Apr 2018 17:17:03	U	SMPL
J2512-03MSD - 1	TP-3-2MSD	16702	3.9752	-	7471B	PPB	24 Apr 2018 17:21:59	U	SMPL
J2512-04 - 1	TP-5-7-9	562	0.1112	-	7471B	PPB	24 Apr 2018 17:24:16	U	SMPL
J2522-01 - 1	CO-W-1	3046	0.7059	-	7471B	PPB	24 Apr 2018 17:26:33	U	SMPL
J2522-02 - 1	CO-S-1	1496	0.3348	-	7471B	PPB	24 Apr 2018 17:28:51	U	SMPL
J2522-03 - 1	CO-W-2	6666	1.5725	-	7471B	PPB	24 Apr 2018 17:31:07	U	SMPL
CCV19 - 1	CCV19	21212	5.0549	-	7471B	PPB	24 Apr 2018 17:33:24	U	SMPL
CCB19 - 1	CCB19	-4	-0.0243	-	7471B	PPB	24 Apr 2018 17:35:40	U	SMPL
J2522-04 - 1	CO-S-2	4814	1.1291	-	7471B	PPB	24 Apr 2018 17:37:57	U	SMPL
J2522-05 - 1	CO-W-3	3234	0.7509	-	7471B	PPB	24 Apr 2018 17:40:14	U	SMPL
J2522-06 - 1	CO-S-3	783	0.1641	-	7471B	PPB	24 Apr 2018 17:42:30	U	SMPL
PB108527BL - 1	PBS	18	-0.0190	-	7471B	PPB	24 Apr 2018 17:44:46	U	SMPL
PB108527BS - 1	LCSS	17609	4.1923	-	7471B	PPB	24 Apr 2018 17:47:03	U	SMPL
J2520-01 - 1	TP-8_0.0-0.5	16381	3.8983	-	7471B	PPB	24 Apr 2018 17:49:20	U	SMPL
J2520-02 - 1	TP-8_2.0-2.5	6221	1.4660	-	7471B	PPB	24 Apr 2018 17:51:37	U	SMPL
J2520-03 - 1	TP-8_4.0-4.5	10243	2.4289	-	7471B	PPB	24 Apr 2018 17:53:55	U	SMPL
J2520-04 - 1	TP-7_0.0-0.5	36576	8.7331	-	7471B	PPB	24 Apr 2018 17:56:13	U	SMPL
J2520-05 - 1	TP-7_2.0-2.5	23399	5.5785	-	7471B	PPB	24 Apr 2018 17:58:29	U	SMPL
CCV20 - 1	CCV20	21034	5.0123	-	7471B	PPB	24 Apr 2018 18:00:47	U	SMPL
CCB20 - 1	CCB20	-14	-0.0267	-	7471B	PPB	24 Apr 2018 18:03:04	U	SMPL
J2520-06 - 1	TP-7_4.0-4.5	11436	2.7145	-	7471B	PPB	24 Apr 2018 18:05:22	U	SMPL
J2520-07 - 1	TP-9_0.0-0.5	9195	2.1780	-	7471B	PPB	24 Apr 2018 18:07:38	U	SMPL
J2520-08 - 1	TP-9_2.0-2.5	799	0.1679	-	7471B	PPB	24 Apr 2018 18:09:55	U	SMPL
J2520-09 - 1	TP-9_4.0-4.5	107860	25.7989	-	7471B	PPB	24 Apr 2018 18:12:12	U	SMPL
J2520-10 - 1	PILES_042318	5980	1.4083	-	7471B	PPB	24 Apr 2018 18:14:29	U	SMPL
J2520-11 - 1	TP-6_0.0-0.5	6442	1.5189	-	7471B	PPB	24 Apr 2018 18:16:49	U	SMPL
J2520-11DUP - 1	TP-6_0.0-0.5DUP	6982	1.6482	-	7471B	PPB	24 Apr 2018 18:19:06	U	SMPL
J2520-11MS - 1	TP-6_0.0-0.5MS	21235	5.0604	-	7471B	PPB	24 Apr 2018 18:21:24	U	SMPL
J2520-11MSD - 1	TP-6_0.0-0.5MSD	21500	5.1239	-	7471B	PPB	24 Apr 2018 18:23:41	U	SMPL
J2520-12 - 1	TP-6_2.0-2.5	2758	0.6369	-	7471B	PPB	24 Apr 2018 18:25:59	U	SMPL
CCV21 - 1	CCV21	20872	4.9735	-	7471B	PPB	24 Apr 2018 18:28:16	U	SMPL