



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB94863

Slope : 97.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/24/2018	17:00
2	CAL2	1	Water			21.2	6.99	04/24/2018	17:01
3	CAL3	1	Water			21.2	10.02	04/24/2018	17:02
4	ICV	1	Water			21.3	7.01	04/24/2018	17:03
5	CCV1	1	Water			21.2	2.01	04/24/2018	17:05
6	J2510-03	1	Water			20.6	7.26	04/24/2018	17:06
7	J2543-02	1	Water			20.1	6.31	04/24/2018	17:08
8	J2543-02DUP	1	Water			20.1	6.32	04/24/2018	17:09
9	CCV2	1	Water			21.3	12.01	04/24/2018	17:10

18 94863

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J2510 WorkList ID : 111204 Date : 4/24/2018 4:46:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/23/2018	Water	J2510-03-18	Corrosivity	Cool 4 deg C	PSEG02	G62	FADW4710L	04/23/2018	9040C
04/24/2018	Water	J2543-02-18	Corrosivity	Cool 4 deg C	PSEG01	G61	TOTE	04/24/2018	9040C

Date/Time 04/24/18 11:50 AM
Received by: JD
Relinquished by: JD

Date/Time 04/24/18 6:30 PM
Received by: CP
Relinquished by: JD