



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94865

Slope : 97.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	04/24/2018	17:00
2	CAL2	1	Water			21.2	6.99	04/24/2018	17:01
3	CAL3	1	Water			21.2	10.02	04/24/2018	17:02
4	ICV	1	Water			21.3	7.01	04/24/2018	17:03
5	CCV1	1	Water			21.2	2.01	04/24/2018	17:20
6	J2155-08	1	Solid	20.03	20	22.5	9.42	04/24/2018	17:21
7	J2155-08DUP	1	Solid	20.05	20	22.4	9.43	04/24/2018	17:22
8	J2155-10	1	Solid	20.02	20	22.6	5.98	04/24/2018	17:24
9	J2155-12	1	Solid	20.01	20	22.4	9.61	04/24/2018	17:25
10	J2156-04	1	Solid	20.03	20	22.3	8.52	04/24/2018	17:26
11	J2158-04	1	Solid	20.04	20	22.3	9.13	04/24/2018	17:28
12	J2541-02	1	Solid	20.05	20	21.5	5.85	04/24/2018	17:30
13	CCV2	1	Water			21.1	12.02	04/24/2018	17:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY J2543

WorkList ID : 111203

Date : 4/24/2018 3:23:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2018	Solid	J2155-08	Corrosivity	Cool 4 deg C	PSEG05		NB-08-042418-A	04/24/2018	9045C
04/24/2018	Solid	J2155-10	Corrosivity	Cool 4 deg C	PSEG05		NB-08-042418-B	04/24/2018	9045C
04/24/2018	Solid	J2155-12	Corrosivity	Cool 4 deg C	PSEG05		NB-08-042418-C	04/24/2018	9045C
04/24/2018	Solid	J2156-04	Corrosivity	Cool 4 deg C	PSEG05		HD-02-042418	04/24/2018	9045C
04/24/2018	Solid	J2158-04	Corrosivity	Cool 4 deg C	PSEG05		OK-01-042418	04/24/2018	9045C
04/23/2018	Solid	J2541-02	Corrosivity	Cool 4 deg C	DIST01	G22	TP-3-4-5-WC-1	04/23/2018	9045C

Date/Time

Received by:

Relinquished by:

04/24/18 4:50 PM

CP

CP

Date/Time

Received by:

Relinquished by:

04/24/18 4:30 PM

CP

CP