



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94877

Slope : 97.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2271         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2220         |
| buffer solution pH 7 yellow    | W2219         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 21.2             | 4.01        | 04/24/2018 | 17:00     |
| 2   | CAL2        | 1  | Water  |             |             | 21.2             | 6.99        | 04/24/2018 | 17:01     |
| 3   | CAL3        | 1  | Water  |             |             | 21.2             | 10.02       | 04/24/2018 | 17:02     |
| 4   | ICV         | 1  | Water  |             |             | 21.3             | 7.01        | 04/24/2018 | 17:03     |
| 5   | CCV1        | 1  | Water  |             |             | 21.2             | 2.01        | 04/24/2018 | 17:05     |
| 6   | J2543-02    | 1  | Water  |             |             | 20.1             | 6.31        | 04/24/2018 | 17:06     |
| 7   | J2543-02DUP | 1  | Water  |             |             | 20.1             | 6.32        | 04/24/2018 | 17:09     |
| 8   | CCV2        | 1  | Water  |             |             | 21.3             | 12.01       | 04/24/2018 | 17:10     |

LB 94002

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY J2510

WorkList ID : 111204

Date : 4/24/2018 4:46:48 PM

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 04/24/2018 | Water  | J2543-02 | Corrosivity | Cool 4 deg C | PSEG01   | G61              | TOTE            | 04/24/2018   | 9040C  |

Date/Time 04/24/18 4:50 PM  
Received by: JP  
Relinquished by: CD

Date/Time 04/24/18 5:35 PM  
Received by: CD  
Relinquished by: JP